

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD041922\
 Data File : VD072632.D
 Acq On : 19 Apr 2022 18:41
 Operator : VA/SY
 Sample : N2447-02
 Misc : 5.06G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VOC

Quant Time: Apr 20 05:04:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D032922S.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 02:44:36 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	423061	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	780852	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	756816	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	345927	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	256709	50.161	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	100.320%
35) Dibromofluoromethane	7.909	113	262226	43.212	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	86.420%
50) Toluene-d8	10.332	98	935736	44.361	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	88.720%
62) 4-Bromofluorobenzene	12.620	95	332041	40.306	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	80.620%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD041922\
Data File : VD072632.D
Acq On : 19 Apr 2022 18:41
Operator : VA/SY
Sample : N2447-02
Misc : 5.06G/5.00ml/MSVOA_D/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VOC

Quant Time: Apr 20 05:04:49 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D032922S.M
Quant Title : SW846 8260
QLast Update : Wed Mar 30 02:44:36 2022
Response via : Initial Calibration

