

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD042021\
 Data File : VD068912.D
 Acq On : 20 Apr 2021 09:34
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 21 08:04:27 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D040721S.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 08 03:50:31 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	64	0.00
2 T	Dichlorodifluoromethane	0.495	0.463	6.5	61	0.00
3 P	Chloromethane	0.582	0.504	13.4	57	0.00
4 C	Vinyl Chloride	0.732	0.690	5.7#	62	0.00
5 T	Bromomethane	0.547	0.508	7.1	65	0.00
6 T	Chloroethane	0.481	0.463	3.7	64	0.00
7 T	Trichlorofluoromethane	0.861	0.893	-3.7	70	0.00
8 T	Diethyl Ether	0.233	0.229	1.7	59	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.520	0.508	2.3	60	0.00
10 T	Methyl Iodide	0.538	0.480	10.8	49#	0.00
11 T	Tert butyl alcohol	0.034	0.025	26.5#	54	0.00
12 CM	1,1-Dichloroethene	0.501	0.471	6.0#	59	0.00
13 T	Acrolein	0.021	0.016	23.8	49#	0.00
14 T	Allyl chloride	0.564	0.461	18.3	51	-0.01
15 T	Acrylonitrile	0.104	0.092	11.5	55	0.00
16 T	Acetone	0.073	0.077	-5.5	67	0.00
17 T	Carbon Disulfide	1.620	1.323	18.3	50#	0.00
18 T	Methyl Acetate	0.201	0.163	18.9	52	0.00
19 T	Methyl tert-butyl Ether	1.110	1.046	5.8	58	0.00
20 T	Methylene Chloride	0.679	0.493	27.4#	49#	0.00
21 T	trans-1,2-Dichloroethene	0.643	0.580	9.8	58	0.00
22 T	Diisopropyl ether	1.367	1.204	11.9	55	0.00
23 T	Vinyl Acetate	0.724	0.677	6.5	56	0.00
24 P	1,1-Dichloroethane	1.030	0.921	10.6	58	0.00
25 T	2-Butanone	0.104	0.099	4.8	58	0.00
26 T	2,2-Dichloropropane	0.842	0.826	1.9	66	0.00
27 T	cis-1,2-Dichloroethene	0.639	0.613	4.1	61	0.00
28 T	Bromochloromethane	0.372	0.311	16.4	53	0.00
29 T	Tetrahydrofuran	0.070	0.058	17.1	50#	0.00
30 C	Chloroform	1.128	0.959	15.0#	55	0.00
31 T	Cyclohexane	0.879	0.731	16.8	55	0.00
32 T	1,1,1-Trichloroethane	0.986	0.879	10.9	57	0.00
33 S	1,2-Dichloroethane-d4	0.528	0.432	18.2	54	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	58	0.00
35 S	Dibromofluoromethane	0.325	0.309	4.9	56	0.00
36 T	1,1-Dichloropropene	0.506	0.524	-3.6	60	0.00
37 T	Ethyl Acetate	0.149	0.140	6.0	57	0.00
38 T	Carbon Tetrachloride	0.501	0.537	-7.2	62	0.00
39 T	Methylcyclohexane	0.590	0.630	-6.8	59	0.00
40 TM	Benzene	1.481	1.366	7.8	54	0.00
41 T	Methacrylonitrile	0.088	0.082	6.8	66	0.00
42 TM	1,2-Dichloroethane	0.391	0.402	-2.8	60	0.00
43 T	Isopropyl Acetate	0.303	0.294	3.0	55	0.00
44 TM	Trichloroethene	0.409	0.412	-0.7	59	0.00
45 C	1,2-Dichloropropane	0.354	0.332	6.2#	55	0.00
46 T	Dibromomethane	0.197	0.198	-0.5	58	0.00
47 T	Bromodichloromethane	0.515	0.525	-1.9	60	0.00
48 T	Methyl methacrylate	0.149	0.126	15.4	46#	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD042021\
 Data File : VD068912.D
 Acq On : 20 Apr 2021 09:34
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 21 08:04:27 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D040721S.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 08 03:50:31 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	59	0.00
50 S	Toluene-d8	1.227	1.282	-4.5	61	0.00
51 T	4-Methyl-2-Pentanone	0.153	0.159	-3.9	57	0.00
52 CM	Toluene	0.938	0.999	-6.5#	60	0.00
53 T	t-1,3-Dichloropropene	0.456	0.473	-3.7	59	0.00
54 T	cis-1,3-Dichloropropene	0.546	0.560	-2.6	58	0.00
55 T	1,1,2-Trichloroethane	0.279	0.286	-2.5	59	0.00
56 T	Ethyl methacrylate	0.287	0.303	-5.6	57	0.00
57 T	1,3-Dichloropropane	0.459	0.467	-1.7	59	0.00
58 T	2-Chloroethyl Vinyl ether	0.140	0.152	-8.6	57	0.00
59 T	2-Hexanone	0.102	0.115	-12.7	60	0.00
60 T	Dibromochloromethane	0.349	0.365	-4.6	61	0.00
61 T	1,2-Dibromoethane	0.271	0.278	-2.6	60	0.00
62 S	4-Bromofluorobenzene	0.407	0.421	-3.4	60	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	62	0.00
64 T	Tetrachloroethene	0.350	0.359	-2.6	65	0.00
65 PM	Chlorobenzene	1.082	1.060	2.0	60	0.00
66 T	1,1,1,2-Tetrachloroethane	0.389	0.388	0.3	62	0.00
67 C	Ethyl Benzene	1.874	1.887	-0.7#	60	0.00
68 T	m/p-Xylenes	0.723	0.683	5.5	57	0.00
69 T	o-Xylene	0.658	0.685	-4.1	62	0.00
70 T	Styrene	1.146	1.159	-1.1	59	0.00
71 P	Bromoform	0.208	0.210	-1.0	62	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	63	0.00
73 T	Isopropylbenzene	3.651	3.715	-1.8	62	0.00
74 T	N-amyl acetate	0.619	0.570	7.9	56	0.00
75 P	1,1,2,2-Tetrachloroethane	0.682	0.629	7.8	59	0.00
76 T	1,2,3-Trichloropropane	0.457	0.407	10.9	57	0.00
77 T	Bromobenzene	0.878	0.860	2.1	63	0.00
78 T	n-propylbenzene	4.434	4.442	-0.2	61	0.00
79 T	2-Chlorotoluene	2.477	2.385	3.7	60	0.00
80 T	1,3,5-Trimethylbenzene	3.052	3.121	-2.3	61	0.00
81 T	trans-1,4-Dichloro-2-butene	0.197	0.193	2.0	58	0.00
82 T	4-Chlorotoluene	2.626	2.543	3.2	61	0.00
83 T	tert-Butylbenzene	2.590	2.631	-1.6	62	0.00
84 T	1,2,4-Trimethylbenzene	3.083	3.127	-1.4	61	0.00
85 T	sec-Butylbenzene	3.654	3.790	-3.7	62	0.00
86 T	p-Isopropyltoluene	3.338	3.497	-4.8	63	0.00
87 T	1,3-Dichlorobenzene	1.739	1.695	2.5	61	0.00
88 T	1,4-Dichlorobenzene	1.721	1.687	2.0	62	0.00
89 T	n-Butylbenzene	3.144	3.273	-4.1	62	0.00
90 T	Hexachloroethane	0.665	0.654	1.7	62	0.00
91 T	1,2-Dichlorobenzene	1.505	1.460	3.0	61	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.104	0.091	12.5	56	0.00
93 T	1,2,4-Trichlorobenzene	0.931	0.952	-2.3	64	0.00
94 T	Hexachlorobutadiene	0.533	0.577	-8.3	65	0.00
95 T	Naphthalene	1.622	1.664	-2.6	60	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD042021\
Data File : VD068912.D
Acq On : 20 Apr 2021 09:34
Operator : VA/SY
Sample : VSTDCCC050
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_D
LabSampleId :
VSTDCCC050

Quant Time: Apr 21 08:04:27 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D040721S.M
Quant Title : SW846 8260
QLast Update : Thu Apr 08 03:50:31 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.814	0.820	-0.7	62	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6