

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD042022\
 Data File : VD072644.D
 Acq On : 20 Apr 2022 13:46
 Operator : VA/SY
 Sample : N2442-05
 Misc : 3.30G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 124042

Quant Time: Apr 20 14:49:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D032922S.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 02:44:36 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	479127	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	868714	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	802948	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	360917	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	255629	44.105	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	88.200%
35) Dibromofluoromethane	7.908	113	266947	39.541	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	79.080%
50) Toluene-d8	10.332	98	1019860	43.459	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	86.920%
62) 4-Bromofluorobenzene	12.620	95	356544	38.903	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	77.800%
Target Compounds						
16) Acetone	4.162	43	15887	14.700	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD042022\
Data File : VD072644.D
Acq On : 20 Apr 2022 13:46
Operator : VA/SY
Sample : N2442-05
Misc : 3.30G/5.00ml/MSVOA_D/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
124042

Quant Time: Apr 20 14:49:43 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D032922S.M
Quant Title : SW846 8260
QLast Update : Wed Mar 30 02:44:36 2022
Response via : Initial Calibration

