

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD042022\
 Data File : VD072658.D
 Acq On : 20 Apr 2022 20:24
 Operator : VA/SY
 Sample : VIBLK
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VIBLK

Quant Time: Apr 21 01:56:34 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D032922S.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 02:44:36 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	590563	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	1037111	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	978919	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	444900	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	289198	40.481	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	80.960%
35) Dibromofluoromethane	7.908	113	308131	38.230	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	76.460%
50) Toluene-d8	10.332	98	1275143	45.514	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	91.020%
62) 4-Bromofluorobenzene	12.620	95	458395	41.895	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	83.780%
Target Compounds						
					Qvalue	
31) Cyclohexane	7.985	56	63020	6.746	ug/l #	82
39) Methylcyclohexane	9.349	83	151232	14.759	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD042022\
Data File : VD072658.D
Acq On : 20 Apr 2022 20:24
Operator : VA/SY
Sample : VIBLK
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VIBLK

Quant Time: Apr 21 01:56:34 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D032922S.M
Quant Title : SW846 8260
QLast Update : Wed Mar 30 02:44:36 2022
Response via : Initial Calibration

