

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD042022\
 Data File : VD072639.D
 Acq On : 20 Apr 2022 11:25
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 20 14:47:02 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D032922S.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 02:44:36 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	81	0.00
2 T	Dichlorodifluoromethane	0.397	0.378	4.8	69	0.00
3 P	Chloromethane	0.367	0.392	-6.8	74	0.00
4 C	Vinyl Chloride	0.385	0.422	-9.6#	74	0.00
5 T	Bromomethane	0.283	0.306	-8.1	78	0.00
6 T	Chloroethane	0.270	0.303	-12.2	78	0.00
7 T	Trichlorofluoromethane	0.861	0.938	-8.9	77	0.00
8 T	Diethyl Ether	0.256	0.280	-9.4	81	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.574	0.617	-7.5	79	0.00
10 T	Methyl Iodide	0.457	0.513	-12.3	73	0.00
11 T	Tert butyl alcohol	0.032	0.050	-56.3#	125	-0.01
12 CM	1,1-Dichloroethene	0.449	0.501	-11.6#	77	0.00
13 T	Acrolein	0.014	0.009	35.7#	78	0.00
14 T	Allyl chloride	0.721	0.800	-11.0	77	0.00
15 T	Acrylonitrile	0.128	0.135	-5.5	81	-0.01
16 T	Acetone	0.113	0.125	-10.6	84	0.00
17 T	Carbon Disulfide	0.795	0.965	-21.4	71	0.00
18 T	Methyl Acetate	0.318	0.292	8.2	82	0.00
19 T	Methyl tert-butyl Ether	1.311	1.424	-8.6	81	0.00
20 T	Methylene Chloride	0.671	0.706	-5.2	85	0.00
21 T	trans-1,2-Dichloroethene	0.502	0.568	-13.1	76	0.00
22 T	Diisopropyl ether	1.804	1.980	-9.8	80	0.00
23 T	Vinyl Acetate	0.871	1.028	-18.0	79	0.00
24 P	1,1-Dichloroethane	1.087	1.184	-8.9	81	0.00
25 T	2-Butanone	0.156	0.169	-8.3	80	0.00
26 T	2,2-Dichloropropane	0.966	1.081	-11.9	83	0.00
27 T	cis-1,2-Dichloroethene	0.682	0.740	-8.5	79	0.00
28 T	Bromochloromethane	0.466	0.505	-8.4	84	0.00
29 T	Tetrahydrofuran	0.094	0.098	-4.3	77	0.00
30 C	Chloroform	1.223	1.303	-6.5#	81	0.00
31 T	Cyclohexane	0.791	0.846	-7.0	75	0.00
32 T	1,1,1-Trichloroethane	1.037	1.153	-11.2	82	0.00
33 S	1,2-Dichloroethane-d4	0.605	0.590	2.5	80	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	80	0.00
35 S	Dibromofluoromethane	0.389	0.414	-6.4	84	0.00
36 T	1,1-Dichloropropene	0.455	0.525	-15.4	77	0.00
37 T	Ethyl Acetate	0.206	0.227	-10.2	80	0.00
38 T	Carbon Tetrachloride	0.535	0.611	-14.2	79	0.00
39 T	Methylcyclohexane	0.498	0.598	-20.1	77	0.00
40 TM	Benzene	1.357	1.569	-15.6	79	0.00
41 T	Methacrylonitrile	0.112	0.136	-21.4	95	0.00
42 TM	1,2-Dichloroethane	0.415	0.455	-9.6	78	0.00
43 T	Isopropyl Acetate	0.395	0.439	-11.1	80	0.00
44 TM	Trichloroethene	0.385	0.441	-14.5	79	0.00
45 C	1,2-Dichloropropane	0.377	0.427	-13.3#	82	0.00
46 T	Dibromomethane	0.198	0.226	-14.1	81	0.00
47 T	Bromodichloromethane	0.569	0.630	-10.7	82	0.00
48 T	Methyl methacrylate	0.192	0.216	-12.5	83	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD042022\
 Data File : VD072639.D
 Acq On : 20 Apr 2022 11:25
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 20 14:47:02 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D032922S.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 02:44:36 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.003	-50.0#	87	0.00
50 S	Toluene-d8	1.351	1.375	-1.8	80	0.00
51 T	4-Methyl-2-Pentanone	0.212	0.236	-11.3	81	0.00
52 CM	Toluene	0.902	1.035	-14.7#	78	0.00
53 T	t-1,3-Dichloropropene	0.496	0.560	-12.9	80	0.00
54 T	cis-1,3-Dichloropropene	0.586	0.673	-14.8	82	0.00
55 T	1,1,2-Trichloroethane	0.300	0.339	-13.0	83	0.00
56 T	Ethyl methacrylate	0.339	0.388	-14.5	78	0.00
57 T	1,3-Dichloropropane	0.496	0.551	-11.1	82	0.00
58 T	2-Chloroethyl Vinyl ether	0.148	0.170	-14.9	83	0.00
59 T	2-Hexanone	0.144	0.166	-15.3	81	0.00
60 T	Dibromochloromethane	0.387	0.428	-10.6	82	0.00
61 T	1,2-Dibromoethane	0.267	0.304	-13.9	82	0.00
62 S	4-Bromofluorobenzene	0.528	0.542	-2.7	84	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	81	0.00
64 T	Tetrachloroethene	0.327	0.360	-10.1	79	0.00
65 PM	Chlorobenzene	1.118	1.218	-8.9	80	0.00
66 T	1,1,1,2-Tetrachloroethane	0.447	0.492	-10.1	83	0.00
67 C	Ethyl Benzene	1.951	2.179	-11.7#	79	0.00
68 T	m/p-Xylenes	0.740	0.841	-13.6	80	0.00
69 T	o-Xylene	0.714	0.793	-11.1	80	0.00
70 T	Styrene	1.258	1.403	-11.5	80	0.00
71 P	Bromoform	0.237	0.251	-5.9	81	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	83	0.00
73 T	Isopropylbenzene	4.038	4.371	-8.2	80	0.00
74 T	N-amyl acetate	0.847	0.887	-4.7	79	0.00
75 P	1,1,2,2-Tetrachloroethane	0.765	0.771	-0.8	83	0.00
76 T	1,2,3-Trichloropropane	0.545	0.568	-4.2	74	0.00
77 T	Bromobenzene	0.905	0.973	-7.5	82	0.00
78 T	n-propylbenzene	4.871	5.273	-8.3	80	0.00
79 T	2-Chlorotoluene	2.849	3.008	-5.6	79	0.00
80 T	1,3,5-Trimethylbenzene	3.458	3.690	-6.7	81	0.00
81 T	trans-1,4-Dichloro-2-butene	0.222	0.245	-10.4	85	0.00
82 T	4-Chlorotoluene	2.967	3.168	-6.8	82	0.00
83 T	tert-Butylbenzene	2.994	3.261	-8.9	82	0.00
84 T	1,2,4-Trimethylbenzene	3.382	3.629	-7.3	82	0.00
85 T	sec-Butylbenzene	4.510	4.837	-7.3	80	0.00
86 T	p-Isopropyltoluene	3.686	3.960	-7.4	81	0.00
87 T	1,3-Dichlorobenzene	1.875	1.984	-5.8	83	0.00
88 T	1,4-Dichlorobenzene	1.890	1.976	-4.6	83	0.00
89 T	n-Butylbenzene	3.512	3.785	-7.8	81	0.00
90 T	Hexachloroethane	0.755	0.781	-3.4	82	0.00
91 T	1,2-Dichlorobenzene	1.654	1.721	-4.1	81	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.113	0.123	-8.8	87	0.00
93 T	1,2,4-Trichlorobenzene	1.025	1.086	-6.0	82	0.00
94 T	Hexachlorobutadiene	0.574	0.614	-7.0	85	0.00
95 T	Naphthalene	1.992	1.968	1.2	78	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD042022\
Data File : VD072639.D
Acq On : 20 Apr 2022 11:25
Operator : VA/SY
Sample : VSTDCCC050
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_D
LabSampleId :
VSTDCCC050

Quant Time: Apr 20 14:47:02 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D032922S.M
Quant Title : SW846 8260
QLast Update : Wed Mar 30 02:44:36 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.895	0.949	-6.0	82	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6