

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD042121\
 Data File : VD068933.D
 Acq On : 21 Apr 2021 18:10
 Operator : VA/SY
 Sample : M2106-13
 Misc : 5.06G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-26-0-2.0

Quant Time: Apr 22 10:23:15 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D040721S.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 08 03:50:31 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.985	168	277185	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	483353	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.644	117	553280	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	284967	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	126504	43.18	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery = 86.36%			
35) Dibromofluoromethane	7.914	113	153542	48.84	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery = 97.68%			
50) Toluene-d8	10.338	98	611502	51.56	ug/l	0.00
Spiked Amount 50.000	Range 49 - 140		Recovery = 103.12%			
62) 4-Bromofluorobenzene	12.632	95	238834	60.63	ug/l	0.00
Spiked Amount 50.000	Range 25 - 144		Recovery = 121.26%			
Target Compounds						
16) Acetone	4.162	43	5620	13.85	ug/l #	83
20) Methylene Chloride	4.979	84	47712	10.74	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD042121\
Data File : VD068933.D
Acq On : 21 Apr 2021 18:10
Operator : VA/SY
Sample : M2106-13
Misc : 5.06G/5.00ml/MSVOA_D/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-26-0-2.0

Quant Time: Apr 22 10:23:15 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D040721S.M
Quant Title : SW846 8260
QLast Update : Thu Apr 08 03:50:31 2021
Response via : Initial Calibration

