

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD042121\
 Data File : VD068935.D
 Acq On : 21 Apr 2021 19:07
 Operator : VA/SY
 Sample : M2106-16
 Misc : 5.08G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-27-0-2.0

Quant Time: Apr 22 10:23:59 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D040721S.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 08 03:50:31 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	294757	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	508858	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.644	117	579889	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	295092	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	126727	40.68	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	81.36%
35) Dibromofluoromethane	7.914	113	163682	49.45	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	98.90%
50) Toluene-d8	10.338	98	648753	51.96	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	103.92%
62) 4-Bromofluorobenzene	12.632	95	256591	61.88	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	123.76%
Target Compounds						
20) Methylene Chloride	4.968	84	39374	7.14	ug/l	Qvalue 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD042121\
Data File : VD068935.D
Acq On : 21 Apr 2021 19:07
Operator : VA/SY
Sample : M2106-16
Misc : 5.08G/5.00ml/MSVOA_D/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-27-0-2.0

Quant Time: Apr 22 10:23:59 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D040721S.M
Quant Title : SW846 8260
QLast Update : Thu Apr 08 03:50:31 2021
Response via : Initial Calibration

