

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD042121\
 Data File : VD068940.D
 Acq On : 21 Apr 2021 21:29
 Operator : VA/SY
 Sample : M2110-02
 Misc : 5.06G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 S-6

Quant Time: Apr 22 10:25:38 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D040721S.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 08 03:50:31 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	271129	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	470537	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.644	117	562105	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	289379	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	123429	43.08	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	86.16%
35) Dibromofluoromethane	7.914	113	152309	49.77	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.54%
50) Toluene-d8	10.338	98	605322	52.43	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	104.86%
62) 4-Bromofluorobenzene	12.632	95	243012	63.37	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	126.74%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD042121\
Data File : VD068940.D
Acq On : 21 Apr 2021 21:29
Operator : VA/SY
Sample : M2110-02
Misc : 5.06G/5.00ml/MSVOA_D/SOIL
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
S-6

Quant Time: Apr 22 10:25:38 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D040721S.M
Quant Title : SW846 8260
QLast Update : Thu Apr 08 03:50:31 2021
Response via : Initial Calibration

