



Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD042122\
 Data File : VD072660.D
 Acq On : 21 Apr 2022 09:05
 Operator : VA/SY
 Sample : BFB
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D032922S.M
 Title : SW846 8260
 Last Update : Wed Mar 30 02:44:36 2022



AutoFind: Scans 1818, 1819, 1820; Background Corrected with Scan 1807

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.7	54941	PASS
75	95	30	60	50.3	148075	PASS
95	95	100	100	100.0	294571	PASS
96	95	5	9	6.5	19101	PASS
173	174	0.00	2	1.1	2375	PASS
174	95	50	100	74.6	219733	PASS
175	174	5	9	7.6	16763	PASS
176	174	95	101	96.5	212139	PASS
177	176	5	9	6.8	14471	PASS