

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD042219\
 Data File : VD061941.D
 Acq On : 22 Apr 2019 16:22
 Operator : VA/AP
 Sample : K2201-09
 Misc : 5.73g/5ml/MSVOA D/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 PL-01-041819-E

Quant Time: Apr 23 00:04:52 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D041519S.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 18 13:49:04 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.04	168	430732	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.21	114	811546	50.00	ug/l	0.00
63) Chlorobenzene-d5	12.37	117	325139	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	14.51	152	219233	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.45	65	190282	52.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.42%	
35) Dibromofluoromethane	6.91	113	294283	48.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.00%	
50) Toluene-d8	10.40	98	124999	8.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	17.20%	
62) 4-Bromofluorobenzene	13.55	95	203053	34.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	68.96%	
Target Compounds						
16) Acetone	3.21	43	14441	16.231	ug/l	97
18) Methyl Acetate	3.63	43	5607	3.144	ug/l #	89
20) Methylene Chloride	3.81	84	31206	2.719	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD042219\
Data File : VD061941.D
Acq On : 22 Apr 2019 16:22
Operator : VA/AP
Sample : K2201-09
Misc : 5.73a/5ml/MSVOA D/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
PL-01-041819-E

Quant Time: Apr 23 00:04:52 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D041519S.M
Quant Title : SW846 8260
QLast Update : Thu Apr 18 13:49:04 2019
Response via : Initial Calibration

