


Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD042220\
Data File : VD065662.D
Acq On : 22 Apr 2020 12:38
Operator : VA/SY
Sample : BFB
Misc : 5.00G/5.00ml/MSVOA D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D042220S.M
Title : SW846 8260
Last Update : Thu Apr 23 04:14:32 2020



AutoFind: Scans 1822, 1823, 1824; Background Corrected with Scan 1814

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.1	8173	PASS
75	95	30	60	49.3	21133	PASS
95	95	100	100	100.0	42826	PASS
96	95	5	9	6.6	2815	PASS
173	174	0.00	2	1.4	559	PASS
174	95	50	100	94.6	40509	PASS
175	174	5	9	8.1	3268	PASS
176	174	95	101	96.8	39232	PASS
177	176	5	9	7.3	2882	PASS