

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD042319\
 Data File : VD061972.D
 Acq On : 23 Apr 2019 16:00
 Operator : VA/AP
 Sample : VD0423SBL01
 Misc : 5.00g/5ml/MSVOA D/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD0423SBL01

Quant Time: Apr 23 16:57:33 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D042319S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 23 14:52:55 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.04	168	673321	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.22	114	999691	50.00	ug/l	0.00
63) Chlorobenzene-d5	12.37	117	827846	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	14.52	152	412762	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.46	65	264610	50.21	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.42%	
35) Dibromofluoromethane	6.92	113	399130	57.73	ug/l	0.00
Spiked Amount	50.000		Recovery	=	115.46%	
50) Toluene-d8	10.41	98	855621	52.01	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.02%	
62) 4-Bromofluorobenzene	13.56	95	374692	56.78	ug/l	0.00
Spiked Amount	50.000		Recovery	=	113.56%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD042319\
Data File : VD061972.D
Acq On : 23 Apr 2019 16:00
Operator : VA/AP
Sample : VD0423SBL01
Misc : 5.00µl/5ml/MSVOA D/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD0423SBL01

Quant Time: Apr 23 16:57:33 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D042319S.M
Quant Title : SW846 8260
QLast Update : Tue Apr 23 14:52:55 2019
Response via : Initial Calibration

