

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD042321\
 Data File : VD068966.D
 Acq On : 23 Apr 2021 14:36
 Operator : VA/SY
 Sample : M2107-20
 Misc : 5.05G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-35-0-2.0

Quant Time: Apr 26 07:55:41 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D040721S.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 08 03:50:31 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	356947	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	605906	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.644	117	729972	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	394460	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	141238	37.44	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	74.88%
35) Dibromofluoromethane	7.914	113	176140	44.69	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	89.38%
50) Toluene-d8	10.338	98	792823	53.33	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	106.66%
62) 4-Bromofluorobenzene	12.632	95	347881	70.45	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	140.90%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD042321\
Data File : VD068966.D
Acq On : 23 Apr 2021 14:36
Operator : VA/SY
Sample : M2107-20
Misc : 5.05G/5.00ml/MSVOA_D/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-35-0-2.0

Quant Time: Apr 26 07:55:41 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D040721S.M
Quant Title : SW846 8260
QLast Update : Thu Apr 08 03:50:31 2021
Response via : Initial Calibration

