



Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD042321\
Data File : VD068957.D
Acq On : 23 Apr 2021 09:22
Operator : VA/SY
Sample : BFB
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D040721S.M
Title : SW846 8260
Last Update : Thu Apr 08 03:50:31 2021



AutoFind: Scans 1820, 1821, 1822; Background Corrected with Scan 1812

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.7	17957	PASS
75	95	30	60	49.9	53640	PASS
95	95	100	100	100.0	107541	PASS
96	95	5	9	6.8	7262	PASS
173	174	0.00	2	0.7	698	PASS
174	95	50	100	88.6	95272	PASS
175	174	5	9	7.3	6992	PASS
176	174	95	101	96.1	91525	PASS
177	176	5	9	6.6	6003	PASS