

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD042419\
 Data File : VD062001.D
 Acq On : 24 Apr 2019 16:51
 Operator : VA/AP
 Sample : K2540-02RE
 Misc : 4.70u/5ml/MSVOA D/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 FK-SF-03-007-ARE

Quant Time: Apr 25 09:43:28 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D042319S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 23 14:52:55 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.05	168	383349	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.22	114	648001	50.00	ug/l	0.00
63) Chlorobenzene-d5	12.37	117	523213	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	14.52	152	176296	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	7.46	65	157504	52.50	ug/l	0.00
Spiked Amount			Recovery	=	105.00%	
35) Dibromofluoromethane	7.05	113	181	0.04	ug/l	0.12
Spiked Amount			Recovery	=	0.08%	
50) Toluene-d8	10.42	98	464207	43.53	ug/l	0.00
Spiked Amount			Recovery	=	87.06%	
62) 4-Bromofluorobenzene	13.56	95	174263	40.74	ug/l	0.00
Spiked Amount			Recovery	=	81.48%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.22	43	79407	110.844	ug/l	96
20) Methylene Chloride	3.82	84	18497	2.259	ug/l	94
52) Toluene	10.52	92	29654	3.721	ug/l	93
68) m/p-Xylenes	12.66	106	26375	4.522	ug/l	94
69) o-Xylene	13.06	106	10893	2.041	ug/l	91
80) 1,3,5-Trimethylbenzene	13.93	105	14887	1.588	ug/l	92
84) 1,2,4-Trimethylbenzene	14.24	105	51723	5.777	ug/l	92
95) Naphthalene	16.13	128	23489	5.533	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD042419\
Data File : VD062001.D
Acq On : 24 Apr 2019 16:51
Operator : VA/AP
Sample : K2540-02RE
Misc : 4.70µl/5ml/MSVOA D/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
FK-SF-03-007-ARE

Quant Time: Apr 25 09:43:28 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D042319S.M
Quant Title : SW846 8260
QLast Update : Tue Apr 23 14:52:55 2019
Response via : Initial Calibration

