

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD042419\
 Data File : VD062004.D
 Acq On : 24 Apr 2019 18:21
 Operator : VA/AP
 Sample : K2539-04RE
 Misc : 4.34g/5ml/MSVOA D/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 FK-SF-03-006-BRE

Quant Time: Apr 25 05:00:27 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D042319S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 23 14:52:55 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.06	168	35560	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.22	114	65795	50.00	ug/l	0.00
63) Chlorobenzene-d5	12.37	117	58827	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	14.52	152	21035	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.46	65	30111	108.19	ug/l	0.00
Spiked Amount	50.000		Recovery	=	216.38%	
35) Dibromofluoromethane	0.00	113	0	0.00	ug/l	
Spiked Amount	50.000		Recovery	=	0.00%	
50) Toluene-d8	10.41	98	47646	44.00	ug/l	0.00
Spiked Amount	50.000		Recovery	=	88.00%	
62) 4-Bromofluorobenzene	13.55	95	22582	52.00	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.00%	
Target Compounds						
					Qvalue	
16) Acetone	3.22	43	19355	291.259	ug/l #	85
52) Toluene	10.52	92	4613	5.701	ug/l	79
68) m/p-Xylenes	12.67	106	1671	2.548	ug/l	90
69) o-Xylene	13.05	106	794	1.323	ug/l	90
80) 1,3,5-Trimethylbenzene	13.94	105	1433	1.281	ug/l	83
84) 1,2,4-Trimethylbenzene	14.24	105	6617	6.194	ug/l	91
95) Naphthalene	16.12	128	6323	12.482	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD042419\
Data File : VD062004.D
Acq On : 24 Apr 2019 18:21
Operator : VA/AP
Sample : K2539-04RE
Misc : 4.34g/5ml/MSVOA D/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
FK-SF-03-006-BRE

Quant Time: Apr 25 05:00:27 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D042319S.M
Quant Title : SW846 8260
QLast Update : Tue Apr 23 14:52:55 2019
Response via : Initial Calibration

