

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD042419\
 Data File : VD062011.D
 Acq On : 24 Apr 2019 21:49
 Operator : VA/AP
 Sample : K2529-01
 Misc : 6.60g/5ml/MSVOA D/SOIL
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 P3P4042219

Quant Time: Apr 25 01:37:14 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D042319S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 23 14:52:55 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	0.00	168	0	0.00	ug/l	-7.05
34) 1,4-Difluorobenzene	8.23	114	187	50.00	ug/l	0.00
63) Chlorobenzene-d5	12.37	117	217	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	0.00	152	0	0.00	ug/l	-14.52

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	0.00	65	0	0.00	ug/l	
Spiked Amount						
			Recovery	=		0.00%
35) Dibromofluoromethane	0.00	113	0	0.00	ug/l	
Spiked Amount						
			Recovery	=		0.00%
50) Toluene-d8	0.00	98	0	0.00	ug/l	
Spiked Amount						
			Recovery	=		0.00%
62) 4-Bromofluorobenzene	0.00	95	0	0.00	ug/l	
Spiked Amount						
			Recovery	=		0.00%

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD042419\
Data File : VD062011.D
Acq On : 24 Apr 2019 21:49
Operator : VA/AP
Sample : K2529-01
Misc : 6.60g/5ml/MSVOA D/SOIL
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
P3P4042219

Quant Time: Apr 25 01:37:14 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D042319S.M
Quant Title : SW846 8260
QLast Update : Tue Apr 23 14:52:55 2019
Response via : Initial Calibration

