

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD042419\
 Data File : VD062023.D
 Acq On : 25 Apr 2019 6:20
 Operator : VA/AP
 Sample : K2200-09
 Misc : 6.56g/5ml/MSVOA D/SOIL
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 EO-02-042319-B

Quant Time: Apr 25 07:38:02 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D042319S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 23 14:52:55 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.06	168	473591	50.00	ug/l	0.01
34) 1,4-Difluorobenzene	8.23	114	739135	50.00	ug/l	0.01
63) Chlorobenzene-d5	12.37	117	397227	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	14.52	152	224413	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.47	65	198180	53.47	ug/l	0.01
Spiked Amount 50.000			Recovery	=	106.94%	
35) Dibromofluoromethane	6.93	113	274498	53.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.40%	
50) Toluene-d8	10.42	98	281368	23.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	46.26%	
62) 4-Bromofluorobenzene	13.56	95	236856	48.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.10%	
Target Compounds						
16) Acetone	3.22	43	12929	14.609	ug/l	95
20) Methylene Chloride	3.83	84	38603	6.164	ug/l	96
67) Ethyl Benzene	12.53	91	13553	1.106	ug/l	100
69) o-Xylene	13.06	106	50863	12.554	ug/l	95
80) 1,3,5-Trimethylbenzene	13.93	105	250564	20.991	ug/l	96
84) 1,2,4-Trimethylbenzene	14.24	105	15130	1.327	ug/l	91
86) p-Isopropyltoluene	14.50	119	47823	3.576	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD042419\
Data File : VD062023.D
Acq On : 25 Apr 2019 6:20
Operator : VA/AP
Sample : K2200-09
Misc : 6.56g/5ml/MSVOA D/SOIL
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
EO-02-042319-B

Quant Time: Apr 25 07:38:02 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D042319S.M
Quant Title : SW846 8260
QLast Update : Tue Apr 23 14:52:55 2019
Response via : Initial Calibration

