

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD042423\
 Data File : VD075789.D
 Acq On : 24 Apr 2023 12:03
 Operator : KP/SY
 Sample : VD0424SBL01
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD0424SBL01

Quant Time: Apr 24 16:08:04 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D041323S.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 13 14:16:29 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	197348	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	358950	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	318990	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	136585	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	100426	53.612	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	107.220%
35) Dibromofluoromethane	7.804	113	122620	50.789	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	101.580%
50) Toluene-d8	10.269	98	452129	46.764	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	93.520%
62) 4-Bromofluorobenzene	12.575	95	112552	43.798	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	87.600%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD042423\
Data File : VD075789.D
Acq On : 24 Apr 2023 12:03
Operator : KP/SY
Sample : VD0424SBL01
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD0424SBL01

Quant Time: Apr 24 16:08:04 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D041323S.M
Quant Title : SW846 8260
QLast Update : Thu Apr 13 14:16:29 2023
Response via : Initial Calibration

