

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD042423\
 Data File : VD075793.D
 Acq On : 24 Apr 2023 13:59
 Operator : KP/SY
 Sample : 02416-05
 Misc : 4.74G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 4-GRAB

Quant Time: Apr 25 06:09:07 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D041323S.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 13 14:16:29 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.881	168	247448	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	460901	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.587	117	351161	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	96088	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	118161	50.308	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 100.620%	
35) Dibromofluoromethane	7.804	113	152214	49.101	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 98.200%	
50) Toluene-d8	10.269	98	530478	42.731	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 85.460%	
62) 4-Bromofluorobenzene	12.575	95	102425	31.041	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 62.080%	
Target Compounds						
20) Methylene Chloride	4.799	84	35552	6.685	ug/l	Qvalue # 77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD042423\
Data File : VD075793.D
Acq On : 24 Apr 2023 13:59
Operator : KP/SY
Sample : 02416-05
Misc : 4.74G/5.00ml/MSVOA_D/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
4-GRAB

Quant Time: Apr 25 06:09:07 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D041323S.M
Quant Title : SW846 8260
QLast Update : Thu Apr 13 14:16:29 2023
Response via : Initial Calibration

