

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD042423\
 Data File : VD075795.D
 Acq On : 24 Apr 2023 14:57
 Operator : KP/SY
 Sample : 02416-03
 Misc : 5.12G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 2-GRAB

Quant Time: Apr 25 06:09:52 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D041323S.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 13 14:16:29 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	259532	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	505522	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.587	117	422541	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	143253	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	130986	53.172	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	=	106.340%	
35) Dibromofluoromethane	7.810	113	144894	42.614	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	=	85.220%	
50) Toluene-d8	10.269	98	588632	43.230	ug/l	0.00
Spiked Amount	50.000	Range 58 - 134	Recovery	=	86.460%	
62) 4-Bromofluorobenzene	12.575	95	134296	37.107	ug/l	0.00
Spiked Amount	50.000	Range 30 - 143	Recovery	=	74.220%	
Target Compounds						
						Qvalue
16) Acetone	4.022	43	16389	36.315	ug/l	# 79
17) Carbon Disulfide	4.275	76	13649	1.435	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD042423\
Data File : VD075795.D
Acq On : 24 Apr 2023 14:57
Operator : KP/SY
Sample : 02416-03
Misc : 5.12G/5.00ml/MSVOA_D/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
2-GRAB

Quant Time: Apr 25 06:09:52 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D041323S.M
Quant Title : SW846 8260
QLast Update : Thu Apr 13 14:16:29 2023
Response via : Initial Calibration

