

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD042423\
 Data File : VD075800.D
 Acq On : 24 Apr 2023 18:19
 Operator : KP/SY
 Sample : 02464-03
 Misc : 5.01G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 MH-92

Quant Time: Apr 25 06:11:21 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D041323S.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 13 14:16:29 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	198700	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	377704	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	326727	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	121188	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	101915	54.036	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	108.080%
35) Dibromofluoromethane	7.810	113	127960	50.369	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	100.740%
50) Toluene-d8	10.269	98	469368	46.137	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	92.280%
62) 4-Bromofluorobenzene	12.575	95	105093	38.865	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	77.720%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD042423\
Data File : VD075800.D
Acq On : 24 Apr 2023 18:19
Operator : KP/SY
Sample : 02464-03
Misc : 5.01G/5.00ml/MSVOA_D/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
MH-92

Quant Time: Apr 25 06:11:21 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D041323S.M
Quant Title : SW846 8260
QLast Update : Thu Apr 13 14:16:29 2023
Response via : Initial Calibration

