

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD042423\
 Data File : VD075802.D
 Acq On : 24 Apr 2023 19:17
 Operator : KP/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 25 06:11:59 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D041323S.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 13 14:16:29 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	96	0.00
2 T	Dichlorodifluoromethane	0.465	0.417	10.3	90	0.00
3 P	Chloromethane	0.839	0.670	20.1	80	0.00
4 C	Vinyl Chloride	1.113	0.864	22.4#	78	0.00
5 T	Bromomethane	0.869	0.664	23.6	78	0.00
6 T	Chloroethane	0.749	0.608	18.8	81	0.00
7 T	Trichlorofluoromethane	0.863	0.796	7.8	92	0.00
8 T	Diethyl Ether	0.268	0.251	6.3	95	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.541	0.541	0.0	100	0.00
10 T	Methyl Iodide	0.740	0.717	3.1	90	0.00
11 T	Tert butyl alcohol	0.024	0.022	8.3	94	0.00
12 CM	1,1-Dichloroethene	0.560	0.538	3.9#	94	0.00
13 T	Acrolein	0.048	0.029	39.6#	59	0.00
14 T	Allyl chloride	0.510	0.489	4.1	92	0.00
15 T	Acrylonitrile	0.100	0.104	-4.0	102	0.00
16 T	Acetone	0.087	0.085	2.3	88	0.00
17 T	Carbon Disulfide	1.833	1.489	18.8	81	0.00
18 T	Methyl Acetate	0.293	0.308	-5.1	103	0.00
19 T	Methyl tert-butyl Ether	1.087	1.078	0.8	94	0.00
20 T	Methylene Chloride	0.800	0.761	4.9	109	0.00
21 T	trans-1,2-Dichloroethene	0.667	0.631	5.4	92	0.00
22 T	Diisopropyl ether	1.153	1.225	-6.2	100	0.00
23 T	Vinyl Acetate	0.535	0.571	-6.7	97	0.00
24 P	1,1-Dichloroethane	0.955	0.945	1.0	98	0.00
25 T	2-Butanone	0.112	0.109	2.7	92	0.00
26 T	2,2-Dichloropropane	0.866	0.778	10.2	91	0.00
27 T	cis-1,2-Dichloroethene	0.739	0.711	3.8	93	0.00
28 T	Bromochloromethane	0.302	0.301	0.3	101	0.00
29 T	Tetrahydrofuran	0.061	0.065	-6.6	100	0.00
30 C	Chloroform	1.083	1.051	3.0#	97	0.00
31 T	Cyclohexane	0.745	0.666	10.6	90	0.00
32 T	1,1,1-Trichloroethane	0.883	0.853	3.4	93	0.00
33 S	1,2-Dichloroethane-d4	0.475	0.442	6.9	94	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	95	0.00
35 S	Dibromofluoromethane	0.336	0.340	-1.2	97	0.00
36 T	1,1-Dichloropropene	0.464	0.450	3.0	90	0.00
37 T	Ethyl Acetate	0.141	0.143	-1.4	98	0.00
38 T	Carbon Tetrachloride	0.409	0.416	-1.7	94	0.00
39 T	Methylcyclohexane	0.536	0.535	0.2	89	0.00
40 TM	Benzene	1.442	1.430	0.8	94	0.00
41 T	Methacrylonitrile	0.068	0.073	-7.4	109	0.00
42 TM	1,2-Dichloroethane	0.328	0.331	-0.9	94	0.00
43 T	Isopropyl Acetate	0.257	0.268	-4.3	96	0.00
44 TM	Trichloroethene	0.408	0.412	-1.0	95	0.00
45 C	1,2-Dichloropropane	0.323	0.338	-4.6#	99	0.00
46 T	Dibromomethane	0.205	0.202	1.5	93	0.00
47 T	Bromodichloromethane	0.469	0.477	-1.7	96	0.00
48 T	Methyl methacrylate	0.115	0.119	-3.5	91	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD042423\
 Data File : VD075802.D
 Acq On : 24 Apr 2023 19:17
 Operator : KP/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 25 06:11:59 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D041323S.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 13 14:16:29 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	93	0.00
50 S	Toluene-d8	1.347	1.317	2.2	83	0.00
51 T	4-Methyl-2-Pentanone	0.137	0.142	-3.6	87	0.00
52 CM	Toluene	0.959	0.942	1.8#	82	0.00
53 T	t-1,3-Dichloropropene	0.450	0.431	4.2	83	0.00
54 T	cis-1,3-Dichloropropene	0.534	0.546	-2.2	87	0.00
55 T	1,1,2-Trichloroethane	0.297	0.292	1.7	88	0.00
56 T	Ethyl methacrylate	0.292	0.297	-1.7	83	0.00
57 T	1,3-Dichloropropane	0.451	0.447	0.9	89	0.00
58 T	2-Chloroethyl Vinyl ether	0.125	0.133	-6.4	89	0.00
59 T	2-Hexanone	0.097	0.101	-4.1	87	0.00
60 T	Dibromochloromethane	0.347	0.336	3.2	85	0.00
61 T	1,2-Dibromoethane	0.283	0.268	5.3	84	0.00
62 S	4-Bromofluorobenzene	0.358	0.362	-1.1	91	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	95	0.00
64 T	Tetrachloroethene	0.360	0.364	-1.1	82	0.00
65 PM	Chlorobenzene	1.071	1.110	-3.6	97	0.00
66 T	1,1,1,2-Tetrachloroethane	0.368	0.377	-2.4	95	0.00
67 C	Ethyl Benzene	1.738	1.821	-4.8#	91	0.00
68 T	m/p-Xylenes	0.712	0.760	-6.7	93	0.00
69 T	o-Xylene	0.644	0.700	-8.7	92	0.00
70 T	Styrene	1.097	1.220	-11.2	92	0.00
71 P	Bromoform	0.192	0.212	-10.4	98	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	92	0.00
73 T	Isopropylbenzene	3.281	3.517	-7.2	93	0.00
74 T	N-amyl acetate	0.482	0.527	-9.3	97	0.00
75 P	1,1,2,2-Tetrachloroethane	0.637	0.662	-3.9	97	0.00
76 T	1,2,3-Trichloropropane	0.408	0.360	11.8	93	0.00
77 T	Bromobenzene	0.846	0.901	-6.5	99	0.00
78 T	n-propylbenzene	4.022	4.230	-5.2	92	0.00
79 T	2-Chlorotoluene	2.281	2.299	-0.8	92	0.00
80 T	1,3,5-Trimethylbenzene	2.736	2.903	-6.1	93	0.00
81 T	trans-1,4-Dichloro-2-butene	0.188	0.204	-8.5	94	0.00
82 T	4-Chlorotoluene	2.372	2.420	-2.0	93	0.00
83 T	tert-Butylbenzene	2.307	2.485	-7.7	94	0.00
84 T	1,2,4-Trimethylbenzene	2.745	2.907	-5.9	92	0.00
85 T	sec-Butylbenzene	3.558	3.799	-6.8	93	0.00
86 T	p-Isopropyltoluene	2.962	3.182	-7.4	93	0.00
87 T	1,3-Dichlorobenzene	1.760	1.799	-2.2	93	0.00
88 T	1,4-Dichlorobenzene	1.739	1.750	-0.6	94	0.00
89 T	n-Butylbenzene	2.729	2.856	-4.7	92	0.00
90 T	Hexachloroethane	0.589	0.584	0.8	94	0.00
91 T	1,2-Dichlorobenzene	1.508	1.558	-3.3	95	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.085	0.081	4.7	91	0.00
93 T	1,2,4-Trichlorobenzene	0.877	0.870	0.8	95	0.00
94 T	Hexachlorobutadiene	0.419	0.422	-0.7	100	0.00
95 T	Naphthalene	1.662	1.723	-3.7	96	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD042423\
Data File : VD075802.D
Acq On : 24 Apr 2023 19:17
Operator : KP/SY
Sample : VSTDCCC050
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_D
LabSampleId :
VSTDCCC050

Quant Time: Apr 25 06:11:59 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D041323S.M
Quant Title : SW846 8260
QLast Update : Thu Apr 13 14:16:29 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.772	0.785	-1.7	99	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6