

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD042519\
 Data File : VD062044.D
 Acq On : 25 Apr 2019 17:51
 Operator : VA/AP
 Sample : K2203-19
 Misc : 6.55g/5ml/MSVOA D/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SU-03-042419-B

Quant Time: Apr 26 04:48:31 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D042319S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 23 14:52:55 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.04	168	79218	50.00	ug/l	-0.01
34) 1,4-Difluorobenzene	8.21	114	128598	50.00	ug/l	-0.01
63) Chlorobenzene-d5	12.36	117	116316	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	14.52	152	55187	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.44	65	60384	97.39	ug/l	-0.01
Spiked Amount 50.000			Recovery	=	194.78%	
35) Dibromofluoromethane	6.92	113	63814	71.75	ug/l	-0.01
Spiked Amount 50.000			Recovery	=	143.50%	
50) Toluene-d8	10.40	98	124515	58.84	ug/l	-0.01
Spiked Amount 50.000			Recovery	=	117.68%	
62) 4-Bromofluorobenzene	13.55	95	57817	68.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	136.22%	
Target Compounds						
16) Acetone	3.21	43	21370	144.354	ug/l	98
20) Methylene Chloride	3.81	84	6913m	6.840	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD042519\
Data File : VD062044.D
Acq On : 25 Apr 2019 17:51
Operator : VA/AP
Sample : K2203-19
Misc : 6.55g/5ml/MSVOA D/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SU-03-042419-B

Quant Time: Apr 26 04:48:31 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D042319S.M
Quant Title : SW846 8260
QLast Update : Tue Apr 23 14:52:55 2019
Response via : Initial Calibration

