

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD042523\
 Data File : VD075818.D
 Acq On : 25 Apr 2023 18:24
 Operator : KP/SY
 Sample : 02419-10
 Misc : 4.09G/10.00ml/MSVOA_D/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 EW925

Quant Time: Apr 26 02:43:57 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\SFAMDLM042023SMA.M
 Quant Title : SFAM01.0
 QLast Update : Wed Apr 26 02:39:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	8.781	114	412996	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.581	117	305609	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.516	152	67143	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.276	65	163175	18.010	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	72.040%
7) Chloroethane-d5	2.799	69	155044	19.806	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	79.240%
11) 1,1-Dichloroethene-d2	3.923	65	36141	17.302	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	69.200%
21) 2-Butanone-d5	6.987	46	45897	58.354	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	116.700%
24) Chloroform-d	7.569	84	235304	22.784	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	91.120%
26) 1,2-Dichloroethane-d4	8.228	65	106957	24.251	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	97.000%
32) Benzene-d6	8.205	84	440513	24.731	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	98.920%
36) 1,2-Dichloropropane-d6	9.210	67	140381	28.473	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	113.880%
41) Toluene-d8	10.269	98	331180	19.860	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	79.440%
43) trans-1,3-Dichloroprop...	10.528	79	50116	24.592	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	98.360%
47) 2-Hexanone-d5	10.875	63	37525	63.212	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	126.420%
56) 1,1,2,2-Tetrachloroeth...	12.651	84	99134	26.277	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	105.120%
66) 1,2-Dichlorobenzene-d4	13.822	152	47861	20.153	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	80.600%
Target Compounds						Qvalue
13) Acetone	4.023	43	11187	16.620	ug/L	73
16) Methylene chloride	4.799	84	25879	3.089	ug/L	93
27) 1,2-Dichloroethane	8.193	62	1865	0.361	ug/L #	68
34) Trichloroethene	8.775	95	10338	2.163	ug/L #	2
39) cis-1,3-Dichloropropene	9.981	75	3281	0.486	ug/L	83
42) Toluene	10.328	91	6224	0.311	ug/L	84
46) Tetrachloroethene	10.816	164	1346	0.343	ug/L #	66
49) Dibromochloromethane	10.816	129	1338	0.340	ug/L #	10
61) 1,2,3-Trichloropropane	13.516	75	8085	7.665	ug/L #	39


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD042523\
Data File : VD075818.D
Acq On : 25 Apr 2023 18:24
Operator : KP/SY
Sample : 02419-10
Misc : 4.09G/10.00ml/MSVOA_D/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
EW925

Quant Time: Apr 26 02:43:57 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\SFAMDLM042023SMA.M
Quant Title : SFAM01.0
QLast Update : Wed Apr 26 02:39:47 2023
Response via : Initial Calibration





#13

Acetone

Concen: 16.620 ug/L

RT: 4.023 min Scan# 41

Delta R.T. 0.006 min

Lab File: VD075818.D

Acq: 25 Apr 2023 18:24

Instrument :

MSVOA_D

ClientSampleId :





EW925

Tgt Ion: 43 Resp: 11187

Ion Ratio Lower Upper

43 100

58 26.1 0.0 87.4



#16

Methylene chloride

Concen: 3.089 ug/L

RT: 4.799 min Scan# 546

Delta R.T. -0.006 min

Lab File: VD075818.D

Acq: 25 Apr 2023 18:24




Tgt Ion: 84 Resp: 25879


Ion Ratio Lower Upper

84 100

86 62.4 45.8 85.2

49 77.7 60.8 112.8









#27
1,2-Dichloroethane
Concen: 0.361 ug/L
RT: 8.193 min Scan# 1146
Delta R.T. -0.135 min
Lab File: VD075818.D
Acq: 25 Apr 2023 18:24




Instrument :
MSVOA_D
ClientSampleId :
EW925


Tgt Ion: 62 Resp: 1865
Ion Ratio Lower Upper
62 100
98 0.0 10.2 15.2#



#34
Trichloroethene
Concen: 2.163 ug/L
RT: 8.775 min Scan# 1222
Delta R.T. -0.253 min
Lab File: VD075818.D
Acq: 25 Apr 2023 18:24

Tgt Ion: 95 Resp: 10338
Ion Ratio Lower Upper
95 100
97 0.0 46.0 85.4#
132 0.0 73.3 136.1#
130 0.0 75.3 139.9#









#39
 cis-1,3-Dichloropropene
 Concen: 0.486 ug/L
 RT: 9.981 min Scan# 1433
 Delta R.T. -0.035 min
 Lab File: VD075818.D
 Acq: 25 Apr 2023 18:24




Instrument :
 MSVOA_D
 ClientSampleId :
 EW925


Tgt Ion: 75 Resp: 3281
 Ion Ratio Lower Upper
 75 100
 77 43.1 23.4 43.6



#42
 Toluene
 Concen: 0.311 ug/L
 RT: 10.328 min Scan# 1486
 Delta R.T. -0.006 min
 Lab File: VD075818.D
 Acq: 25 Apr 2023 18:24

Tgt Ion: 91 Resp: 6224
 Ion Ratio Lower Upper
 91 100
 92 44.3 39.3 72.9









#46
Tetrachloroethene
Concen: 0.343 ug/L
RT: 10.816 min Scan# 1569
Delta R.T. 0.006 min
Lab File: VD075818.D
Acq: 25 Apr 2023 18:24




Instrument :
MSVOA_D
ClientSampleId :
EW925


Tgt Ion:164 Resp: 1346
Ion Ratio Lower Upper
164 100
129 135.3 60.3 112.1#
131 116.6 58.9 109.3#
166 99.0 83.0 154.1



#49
Dibromochloromethane
Concen: 0.340 ug/L
RT: 10.816 min Scan# 1569
Delta R.T. -0.259 min
Lab File: VD075818.D
Acq: 25 Apr 2023 18:24


Tgt Ion:129 Resp: 1338
Ion Ratio Lower Upper
129 100
127 0.0 55.6 103.4#





#61
 1,2,3-Trichloropropane
 Concen: 7.665 ug/L
 RT: 13.516 min Scan# 20
 Delta R.T. 0.788 min
 Lab File: VD075818.D
 Acq: 25 Apr 2023 18:24

Instrument :
 MSVOA_D
 ClientSampleId :
 EW925



Tgt Ion:	75	Resp:	8085
Ion	Ratio	Lower	Upper
75	100		
77	0.8	26.6	39.8#
110	0.9	33.5	50.3#

