

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD042523\
 Data File : VD075819.D
 Acq On : 25 Apr 2023 18:53
 Operator : KP/SY
 Sample : 02419-13
 Misc : 4.32G/10.00ml/MSVOA_D/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 EW926

Quant Time: Apr 26 02:44:13 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\SFAMDLM042023SMA.M
 Quant Title : SFAM01.0
 QLast Update : Wed Apr 26 02:39:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	8.775	114	573615	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.581	117	526628	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.522	152	185488	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.275	65	210946	16.764	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery =	67.040%		
7) Chloroethane-d5	2.805	69	195602	17.990	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery =	71.960%		
11) 1,1-Dichloroethene-d2	3.916	65	50141	17.283	ug/L	0.00
Spiked Amount	25.000	Range 45 - 110	Recovery =	69.120%		
21) 2-Butanone-d5	6.981	46	42321	38.741	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery =	77.480%		
24) Chloroform-d	7.569	84	280211	19.535	ug/L	0.00
Spiked Amount	25.000	Range 40 - 150	Recovery =	78.120%		
26) 1,2-Dichloroethane-d4	8.234	65	118568	19.356	ug/L	0.00
Spiked Amount	25.000	Range 70 - 130	Recovery =	77.440%		
32) Benzene-d6	8.204	84	572092	18.638	ug/L	0.00
Spiked Amount	25.000	Range 20 - 135	Recovery =	74.560%		
36) 1,2-Dichloropropane-d6	9.210	67	170885	20.114	ug/L	0.00
Spiked Amount	25.000	Range 70 - 120	Recovery =	80.440%		
41) Toluene-d8	10.269	98	507444	17.659	ug/L	0.00
Spiked Amount	25.000	Range 30 - 130	Recovery =	70.640%		
43) trans-1,3-Dichloroprop...	10.528	79	64280	18.305	ug/L	0.00
Spiked Amount	25.000	Range 30 - 135	Recovery =	73.200%		
47) 2-Hexanone-d5	10.875	63	38188	37.331	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery =	74.660%		
56) 1,1,2,2-Tetrachloroeth...	12.651	84	118480	18.224	ug/L	0.00
Spiked Amount	25.000	Range 45 - 120	Recovery =	72.880%		
66) 1,2-Dichlorobenzene-d4	13.816	152	132635	20.216	ug/L	0.00
Spiked Amount	25.000	Range 75 - 120	Recovery =	80.880%		
Target Compounds						
13) Acetone	4.022	43	7221	7.724	ug/L	75
16) Methylene chloride	4.799	84	25173	2.163	ug/L	96
25) Chloroform	7.599	83	10270	0.740	ug/L	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD042523\
 Data File : VD075819.D
 Acq On : 25 Apr 2023 18:53
 Operator : KP/SY
 Sample : 02419-13
 Misc : 4.32G/10.00ml/MSVOA_D/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 EW926

Quant Time: Apr 26 02:44:13 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\SFAMDLM042023SMA.M
 Quant Title : SFAM01.0
 QLast Update : Wed Apr 26 02:39:47 2023
 Response via : Initial Calibration

