

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD042621\
 Data File : VD068989.D
 Acq On : 26 Apr 2021 15:33
 Operator : VA/SY
 Sample : M2170-02
 Misc : 8.36G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 DN-1

Quant Time: Apr 27 07:13:03 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D040721S.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 08 03:50:31 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	294223	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	503669	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.643	117	606040	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	305384	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.338	65	127859	41.12	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	82.24%
35) Dibromofluoromethane	7.914	113	161632	49.34	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	98.68%
50) Toluene-d8	10.338	98	645849	52.26	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	104.52%
62) 4-Bromofluorobenzene	12.632	95	284579	69.33	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	138.66%
Target Compounds						
16) Acetone	4.167	43	4360	10.13	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD042621\
Data File : VD068989.D
Acq On : 26 Apr 2021 15:33
Operator : VA/SY
Sample : M2170-02
Misc : 8.36G/5.00ml/MSVOA_D/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
DN-1

Quant Time: Apr 27 07:13:03 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D040721S.M
Quant Title : SW846 8260
QLast Update : Thu Apr 08 03:50:31 2021
Response via : Initial Calibration

