

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD042623\
 Data File : VD075829.D
 Acq On : 26 Apr 2023 12:21
 Operator : KP/SY
 Sample : VIBLK108
 Misc : 5.00G/10.00ml/MSVOA_D/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VIBLK108

Quant Time: Apr 27 01:25:56 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\SFAMDLM042023SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Apr 27 01:23:50 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.781	114	374212	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.581	117	338707	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.522	152	140383	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.275	65	219753	26.769	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	= 107.080%	
7) Chloroethane-d5	2.805	69	178411	25.153	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	= 100.600%	
11) 1,1-Dichloroethene-d2	3.922	65	51765	27.351	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	= 109.400%	
21) 2-Butanone-d5	6.987	46	37508	52.631	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	= 105.260%	
24) Chloroform-d	7.569	84	233737	24.978	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	= 99.920%	
26) 1,2-Dichloroethane-d4	8.234	65	105307	26.352	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	= 105.400%	
32) Benzene-d6	8.204	84	512499	25.961	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	= 103.840%	
36) 1,2-Dichloropropane-d6	9.210	67	141411	25.880	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	= 103.520%	
41) Toluene-d8	10.269	98	470860	25.477	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	= 101.920%	
43) trans-1,3-Dichloroprop...	10.528	79	57964	25.664	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	= 102.640%	
47) 2-Hexanone-d5	10.881	63	31514	47.899	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	= 95.800%	
56) 1,1,2,2-Tetrachloroeth...	12.651	84	105166	25.151	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	= 100.600%	
66) 1,2-Dichlorobenzene-d4	13.816	152	148503	29.907	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	= 119.640%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD042623\
Data File : VD075829.D
Acq On : 26 Apr 2023 12:21
Operator : KP/SY
Sample : VIBLK108
Misc : 5.00G/10.00ml/MSVOA_D/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VIBLK108

Quant Time: Apr 27 01:25:56 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\SFAMDLM042023SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Apr 27 01:23:50 2023
Response via : Initial Calibration

