

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD042623\
 Data File : VD075847.D
 Acq On : 26 Apr 2023 21:04
 Operator : KP/SY
 Sample : 02447-21
 Misc : 4.04G/10.00ml/MSVOA_D/SOIL
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 EW939

Quant Time: Apr 27 02:07:02 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\SFAMDLM042023SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Apr 27 02:02:18 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.781	114	530218	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.581	117	373597	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.522	152	79929	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.275	65	204831	17.610	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	70.440%		
7) Chloroethane-d5	2.799	69	194204	19.324	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	77.280%		
11) 1,1-Dichloroethene-d2	3.922	65	46374	17.293	ug/L	0.00
Spiked Amount 25.000	Range 45 - 110		Recovery =	69.160%		
21) 2-Butanone-d5	6.981	46	45576	45.135	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	90.280%		
24) Chloroform-d	7.569	84	285326	21.519	ug/L	0.00
Spiked Amount 25.000	Range 40 - 150		Recovery =	86.080%		
26) 1,2-Dichloroethane-d4	8.234	65	121631	21.481	ug/L	0.00
Spiked Amount 25.000	Range 70 - 130		Recovery =	85.920%		
32) Benzene-d6	8.204	84	539899	24.795	ug/L	0.00
Spiked Amount 25.000	Range 20 - 135		Recovery =	99.160%		
36) 1,2-Dichloropropane-d6	9.210	67	171395	28.438	ug/L	0.00
Spiked Amount 25.000	Range 70 - 120		Recovery =	113.760%		
41) Toluene-d8	10.269	98	415461	20.380	ug/L	0.00
Spiked Amount 25.000	Range 30 - 130		Recovery =	81.520%		
43) trans-1,3-Dichloroprop...	10.522	79	51667	20.740	ug/L	0.00
Spiked Amount 25.000	Range 30 - 135		Recovery =	82.960%		
47) 2-Hexanone-d5	10.881	63	38306	52.785	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	105.560%		
56) 1,1,2,2-Tetrachloroeth...	12.651	84	108363	23.496	ug/L	0.00
Spiked Amount 25.000	Range 45 - 120		Recovery =	94.000%		
66) 1,2-Dichlorobenzene-d4	13.816	152	55646	19.682	ug/L	0.00
Spiked Amount 25.000	Range 75 - 120		Recovery =	78.720%		
Target Compounds						Qvalue
16) Methylene chloride	4.799	84	18621	1.731	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD042623\
Data File : VD075847.D
Acq On : 26 Apr 2023 21:04
Operator : KP/SY
Sample : 02447-21
Misc : 4.04G/10.00ml/MSVOA_D/SOIL
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
EW939

Quant Time: Apr 27 02:07:02 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\SFAMDLM042023SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Apr 27 02:02:18 2023
Response via : Initial Calibration

