

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD042821\
 Data File : VD068996.D
 Acq On : 28 Apr 2021 11:03
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :

Quant Time: Apr 29 06:28:57 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D040721S.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 08 03:50:31 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	281759	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	436979	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.643	117	405730	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	201834	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	131120	44.03	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	88.06%
35) Dibromofluoromethane	7.914	113	148532	52.26	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	104.52%
50) Toluene-d8	10.338	98	547375	51.05	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	102.10%
62) 4-Bromofluorobenzene	12.632	95	179132	50.30	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	100.60%

Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.991	85	130404	46.71	ug/l	99
3) Chloromethane	2.209	50	137562	41.91	ug/l	98
4) Vinyl Chloride	2.350	62	193765	46.99	ug/l	97
5) Bromomethane	2.762	94	156982	50.93	ug/l	95
6) Chloroethane	2.920	64	139622	51.53	ug/l	98
7) Trichlorofluoromethane	3.273	101	259083	53.40	ug/l	100
8) Diethyl Ether	3.715	74	64829	49.48	ug/l	90
9) 1,1,2-Trichlorotrifluo...	4.103	101	157860	53.91	ug/l	97
10) Methyl Iodide	4.303	142	159636	46.74	ug/l	98
11) Tert butyl alcohol	5.220	59	39263	248.16	ug/l #	90
12) 1,1-Dichloroethene	4.067	96	144364	51.19	ug/l	95
13) Acrolein	3.926	56	25017	213.45	ug/l	98
14) Allyl chloride	4.714	41	143572	45.17	ug/l	92
15) Acrylonitrile	5.420	53	122135	207.69	ug/l	98
16) Acetone	4.162	43	123811	300.25	ug/l	98
17) Carbon Disulfide	4.414	76	410822	45.00	ug/l	99
18) Methyl Acetate	4.726	43	56091	49.51	ug/l	100
19) Methyl tert-butyl Ether	5.485	73	278167	44.45	ug/l	98
20) Methylene Chloride	4.967	84	167423	50.14	ug/l	94
21) trans-1,2-Dichloroethene	5.473	96	158979	43.90	ug/l	95
22) Diisopropyl ether	6.367	45	337671	43.85	ug/l	99
23) Vinyl Acetate	6.308	43	926543	227.02	ug/l	96
24) 1,1-Dichloroethane	6.267	63	263413	45.37	ug/l	98
25) 2-Butanone	7.214	43	145394	249.17	ug/l	99
26) 2,2-Dichloropropane	7.208	77	253359	53.42	ug/l	100
27) cis-1,2-Dichloroethene	7.208	96	186208	51.73	ug/l	97
28) Bromochloromethane	7.550	49	91999	43.90	ug/l	88
29) Tetrahydrofuran	7.567	42	85361	217.52	ug/l	93
30) Chloroform	7.708	83	294304	46.31	ug/l	98
31) Cyclohexane	7.985	56	213132	43.01	ug/l	96
32) 1,1,1-Trichloroethane	7.903	97	271969	48.97	ug/l	99
36) 1,1-Dichloropropene	8.114	75	223616	50.56	ug/l	97
37) Ethyl Acetate	7.291	43	67788	52.18	ug/l	97
38) Carbon Tetrachloride	8.097	117	244652	55.87	ug/l	98
39) Methylcyclohexane	9.355	83	278499	53.99	ug/l	96
40) Benzene	8.350	78	647083	49.99	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD042821\
 Data File : VD068996.D
 Acq On : 28 Apr 2021 11:03
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :

Quant Time: Apr 29 06:28:57 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D040721S.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 08 03:50:31 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.561	41	88780	115.43	ug/l #	45
42) 1,2-Dichloroethane	8.420	62	169979	49.68	ug/l	98
43) Isopropyl Acetate	8.450	43	129985	49.08	ug/l	99
44) Trichloroethene	9.114	130	190674	53.34	ug/l	92
45) 1,2-Dichloropropane	9.385	63	152096	49.20	ug/l	100
46) Dibromomethane	9.473	93	87703	50.93	ug/l	94
47) Bromodichloromethane	9.661	83	228798	50.85	ug/l	98
48) Methyl methacrylate	9.455	41	63524	48.67	ug/l	92
49) 1,4-Dioxane	9.467	88	20139	1104.52	ug/l	90
51) 4-Methyl-2-Pentanone	10.226	43	338721	253.74	ug/l	99
52) Toluene	10.402	92	440033	53.69	ug/l	98
53) t-1,3-Dichloropropene	10.620	75	204878	51.38	ug/l	100
54) cis-1,3-Dichloropropene	10.091	75	243903	51.13	ug/l	97
55) 1,1,2-Trichloroethane	10.796	97	125389	51.49	ug/l	98
56) Ethyl methacrylate	10.661	69	133533	53.27	ug/l	98
57) 1,3-Dichloropropane	10.944	76	199962	49.87	ug/l	98
58) 2-Chloroethyl Vinyl ether	9.944	63	345535	258.32	ug/l	97
59) 2-Hexanone	10.985	43	241582	270.47	ug/l	98
60) Dibromochloromethane	11.138	129	166033	54.38	ug/l	99
61) 1,2-Dibromoethane	11.244	107	120788	51.04	ug/l	99
64) Tetrachloroethene	10.873	164	156654	55.08	ug/l	94
65) Chlorobenzene	11.667	112	471873	53.73	ug/l	100
66) 1,1,1,2-Tetrachloroethane	11.743	131	176846	56.00	ug/l	99
67) Ethyl Benzene	11.743	91	839198	55.18	ug/l	99
68) m/p-Xylenes	11.849	106	648594	110.52	ug/l	98
69) o-Xylene	12.179	106	290228	54.39	ug/l	99
70) Styrene	12.191	104	518440	55.75	ug/l	99
71) Bromoform	12.355	173	92315	54.57	ug/l #	95
73) Isopropylbenzene	12.479	105	828588	56.22	ug/l	100
74) N-amyl acetate	12.285	43	121408	48.58	ug/l	98
75) 1,1,2,2-Tetrachloroethane	12.726	83	140489	51.00	ug/l	99
76) 1,2,3-Trichloropropane	12.779	75	171943	93.25	ug/l #	100
77) Bromobenzene	12.755	156	189034	53.36	ug/l	94
78) n-propylbenzene	12.814	91	985127	55.03	ug/l	99
79) 2-Chlorotoluene	12.902	91	521745	52.18	ug/l	97
80) 1,3,5-Trimethylbenzene	12.955	105	681704	55.33	ug/l	99
81) trans-1,4-Dichloro-2-b...	12.526	75	40582	51.08	ug/l	99
82) 4-Chlorotoluene	13.002	91	555277	52.39	ug/l	97
83) tert-Butylbenzene	13.220	119	579219	55.40	ug/l	99
84) 1,2,4-Trimethylbenzene	13.261	105	681252	54.75	ug/l	98
85) sec-Butylbenzene	13.396	105	829742	56.25	ug/l	99
86) p-Isopropyltoluene	13.508	119	763929	56.70	ug/l	99
87) 1,3-Dichlorobenzene	13.514	146	375602	53.51	ug/l	98
88) 1,4-Dichlorobenzene	13.590	146	367223	52.86	ug/l	99
89) n-Butylbenzene	13.838	91	716178	56.43	ug/l	100
90) Hexachloroethane	14.108	117	142716	53.18	ug/l	95
91) 1,2-Dichlorobenzene	13.885	146	319051	52.53	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.496	75	21342	50.97	ug/l	93
93) 1,2,4-Trichlorobenzene	15.155	180	212197	56.46	ug/l	99
94) Hexachlorobutadiene	15.261	225	127086	59.06	ug/l	99
95) Naphthalene	15.396	128	366431	55.97	ug/l	99
96) 1,2,3-Trichlorobenzene	15.590	180	183626	55.86	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD042821\
Data File : VD068996.D
Acq On : 28 Apr 2021 11:03
Operator : VA/SY
Sample : VSTDCCC050
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 2 Sample Multiplier: 1


Instrument :
MSVOA_D
ClientSampleId :


Quant Time: Apr 29 06:28:57 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D040721S.M
Quant Title : SW846 8260
QLast Update : Thu Apr 08 03:50:31 2021
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Apr 29 06:28:57 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D040721S.M
Quant Title : SW846 8260
QLast Update : Thu Apr 08 03:50:31 2021
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.00 ug/l
RT: 7.979 min Scan# 1030
Delta R.T. -0.000 min
Lab File: VD068996.D
Acq: 28 Apr 2021 11:03




Instrument :
MSVOA_D
ClientSampled :


Tgt Ion:168 Resp: 281759
Ion Ratio Lower Upper
168 100
99 57.4 46.2 69.2



#2
Dichlorodifluoromethane
Concen: 46.71 ug/l
RT: 1.991 min Scan# 12
Delta R.T. 0.000 min
Lab File: VD068996.D
Acq: 28 Apr 2021 11:03

Tgt Ion: 85 Resp: 130404
Ion Ratio Lower Upper
85 100
87 32.1 15.7 47.0









#3
Chloromethane
Concen: 41.91 ug/l
RT: 2.209 min Scan# 49
Delta R.T. -0.000 min
Lab File: VD068996.D
Acq: 28 Apr 2021 11:03




Instrument :
MSVOA_D
ClientSampled :


Tgt Ion: 50 Resp: 137562
Ion Ratio Lower Upper
50 100
52 32.3 26.7 40.1



#4
Vinyl Chloride
Concen: 46.99 ug/l
RT: 2.350 min Scan# 73
Delta R.T. -0.000 min
Lab File: VD068996.D
Acq: 28 Apr 2021 11:03

Tgt Ion: 62 Resp: 193765
Ion Ratio Lower Upper
62 100
64 32.4 27.1 40.7









#5
Bromomethane
Concen: 50.93 ug/l
RT: 2.762 min Scan# 143
Delta R.T. -0.005 min
Lab File: VD068996.D
Acq: 28 Apr 2021 11:03




Instrument :
MSVOA_D
ClientSampled :


Tgt Ion: 94 Resp: 156982
Ion Ratio Lower Upper
94 100
96 88.8 74.7 112.1



#6
Chloroethane
Concen: 51.53 ug/l
RT: 2.920 min Scan# 170
Delta R.T. 0.000 min
Lab File: VD068996.D
Acq: 28 Apr 2021 11:03

Tgt Ion: 64 Resp: 139622
Ion Ratio Lower Upper
64 100
66 31.2 26.0 39.0









#7
Trichlorofluoromethane
Concen: 53.40 ug/l
RT: 3.273 min Scan# 230
Delta R.T. 0.000 min
Lab File: VD068996.D
Acq: 28 Apr 2021 11:03




Instrument :
MSVOA_D
ClientSampled :


Tgt Ion:101 Resp: 259083
Ion Ratio Lower Upper
101 100
103 66.8 53.4 80.2



#8
Diethyl Ether
Concen: 49.48 ug/l
RT: 3.715 min Scan# 305
Delta R.T. 0.006 min
Lab File: VD068996.D
Acq: 28 Apr 2021 11:03

Tgt Ion: 74 Resp: 64829
Ion Ratio Lower Upper
74 100
45 72.5 40.5 121.5









#9
1,1,2-Trichlorotrifluoroethane
Concen: 53.91 ug/l
RT: 4.103 min Scan# 371
Delta R.T. 0.006 min
Lab File: VD068996.D
Acq: 28 Apr 2021 11:03




Instrument :
MSVOA_D
ClientSampled :


Tgt Ion:101 Resp: 157860
Ion Ratio Lower Upper
101 100
85 43.6 33.8 50.6
151 77.8 59.4 89.2



#10
Methyl Iodide
Concen: 46.74 ug/l
RT: 4.303 min Scan# 405
Delta R.T. -0.000 min
Lab File: VD068996.D
Acq: 28 Apr 2021 11:03

Tgt Ion:142 Resp: 159636
Ion Ratio Lower Upper
142 100
127 41.6 34.7 52.1
141 14.0 11.0 16.4









#11

Tert butyl alcohol
Concen: 248.16 ug/l
RT: 5.220 min Scan# 561
Delta R.T. 0.000 min
Lab File: VD068996.D
Acq: 28 Apr 2021 11:03

Instrument :
MSVOA_D
ClientSampled :




Tgt Ion: 59 Resp: 39263
Ion Ratio Lower Upper
59 100
57 6.1 7.8 11.8#



#12

1,1-Dichloroethene
Concen: 51.19 ug/l
RT: 4.067 min Scan# 365
Delta R.T. -0.006 min
Lab File: VD068996.D
Acq: 28 Apr 2021 11:03

Tgt Ion: 96 Resp: 144364
Ion Ratio Lower Upper
96 100
61 134.1 112.2 168.2
98 67.1 50.2 75.2









#13
Acrolein
Concen: 213.45 ug/l
RT: 3.926 min Scan# 341
Delta R.T. 0.000 min
Lab File: VD068996.D
Acq: 28 Apr 2021 11:03




Instrument :
MSVOA_D
ClientSampled :


Tgt Ion: 56 Resp: 25017
Ion Ratio Lower Upper
56 100
55 70.4 54.7 82.1



#14
Allyl chloride
Concen: 45.17 ug/l
RT: 4.714 min Scan# 475
Delta R.T. -0.006 min
Lab File: VD068996.D
Acq: 28 Apr 2021 11:03

Tgt Ion: 41 Resp: 143572
Ion Ratio Lower Upper
41 100
39 81.0 60.6 91.0
76 57.8 40.4 60.6









#15
Acrylonitrile
Concen: 207.69 ug/l
RT: 5.420 min Scan# 595
Delta R.T. -0.006 min
Lab File: VD068996.D
Acq: 28 Apr 2021 11:03




Instrument :
MSVOA_D
ClientSampled :


Tgt Ion: 53	Resp: 122135
Ion Ratio	Lower Upper
53 100	
52 80.9	63.3 94.9
51 37.1	28.3 42.5



#16
Acetone
Concen: 300.25 ug/l
RT: 4.162 min Scan# 381
Delta R.T. 0.001 min
Lab File: VD068996.D
Acq: 28 Apr 2021 11:03

Tgt Ion: 43	Resp: 123811
Ion Ratio	Lower Upper
43 100	
58 36.3	28.0 42.0









#17
Carbon Disulfide
Concen: 45.00 ug/l
RT: 4.414 min Scan# 424
Delta R.T. 0.000 min
Lab File: VD068996.D
Acq: 28 Apr 2021 11:03




Instrument :
MSVOA_D
ClientSampled :


Tgt Ion: 76 Resp: 410822
Ion Ratio Lower Upper
76 100
78 9.5 7.3 10.9



#18
Methyl Acetate
Concen: 49.51 ug/l
RT: 4.726 min Scan# 477
Delta R.T. 0.006 min
Lab File: VD068996.D
Acq: 28 Apr 2021 11:03

Tgt Ion: 43 Resp: 56091
Ion Ratio Lower Upper
43 100
74 32.0 25.4 38.2









#19

Methyl tert-butyl Ether
 Concen: 44.45 ug/l
 RT: 5.485 min Scan# 606
 Delta R.T. 0.000 min
 Lab File: VD068996.D
 Acq: 28 Apr 2021 11:03

Instrument :
 MSVOA_D
 ClientSampleId :




Tgt Ion: 73 Resp: 278167
 Ion Ratio Lower Upper
 73 100
 57 22.3 17.3 25.9



#20

Methylene Chloride
 Concen: 50.14 ug/l
 RT: 4.967 min Scan# 518
 Delta R.T. -0.006 min
 Lab File: VD068996.D
 Acq: 28 Apr 2021 11:03


Tgt Ion: 84 Resp: 167423
 Ion Ratio Lower Upper
 84 100
 49 95.1 83.4 125.0
 51 30.8 23.4 35.2
 86 65.4 51.0 76.4








#21
trans-1,2-Dichloroethene
Concen: 43.90 ug/l
RT: 5.473 min Scan# 604
Delta R.T. 0.000 min
Lab File: VD068996.D
Acq: 28 Apr 2021 11:03

Instrument :
MSVOA_D
ClientSampleId :







Tgt Ion: 96 Resp: 158979
Ion Ratio Lower Upper
96 100
61 111.7 91.8 137.6
98 54.7 49.3 73.9



#22
Diisopropyl ether
Concen: 43.85 ug/l
RT: 6.367 min Scan# 756
Delta R.T. 0.000 min
Lab File: VD068996.D
Acq: 28 Apr 2021 11:03

Tgt Ion: 45 Resp: 337671
Ion Ratio Lower Upper
45 100
43 53.5 41.8 62.8
87 36.9 29.2 43.8
59 13.3 10.3 15.5









#23
 Vinyl Acetate
 Concen: 227.02 ug/l
 RT: 6.308 min Scan# 746
 Delta R.T. 0.000 min
 Lab File: VD068996.D
 Acq: 28 Apr 2021 11:03




Instrument :
 MSVOA_D
 ClientSampled :


Tgt Ion: 43 Resp: 926543
 Ion Ratio Lower Upper
 43 100
 86 15.6 11.0 16.6



#24
 1,1-Dichloroethane
 Concen: 45.37 ug/l
 RT: 6.267 min Scan# 739
 Delta R.T. 0.000 min
 Lab File: VD068996.D
 Acq: 28 Apr 2021 11:03


Tgt Ion: 63 Resp: 263413
 Ion Ratio Lower Upper
 63 100
 98 7.8 3.8 11.2
 100 5.6 2.3 6.9








#25
2-Butanone
Concen: 249.17 ug/l
RT: 7.214 min Scan# 900
Delta R.T. 0.000 min
Lab File: VD068996.D
Acq: 28 Apr 2021 11:03

Instrument :
MSVOA_D
ClientSampleId :







Tgt Ion: 43 Resp: 145394
Ion Ratio Lower Upper
43 100
72 36.3 28.6 43.0



#26
2,2-Dichloropropane
Concen: 53.42 ug/l
RT: 7.208 min Scan# 899
Delta R.T. -0.006 min
Lab File: VD068996.D
Acq: 28 Apr 2021 11:03


Tgt Ion: 77 Resp: 253359
Ion Ratio Lower Upper
77 100
97 24.3 12.0 36.1








#27
 cis-1,2-Dichloroethene
 Concen: 51.73 ug/l
 RT: 7.208 min Scan# 899
 Delta R.T. -0.006 min
 Lab File: VD068996.D
 Acq: 28 Apr 2021 11:03

Instrument :
 MSVOA_D
 ClientSampled :







Tgt Ion: 96 Resp: 186208
 Ion Ratio Lower Upper
 96 100
 61 125.7 0.0 258.8
 98 66.5 0.0 128.4



#28
 Bromochloromethane
 Concen: 43.90 ug/l
 RT: 7.550 min Scan# 957
 Delta R.T. 0.006 min
 Lab File: VD068996.D
 Acq: 28 Apr 2021 11:03

Tgt Ion: 49 Resp: 91999
 Ion Ratio Lower Upper
 49 100
 129 3.5 0.0 5.2
 130 105.0 74.6 112.0





#29

Tetrahydrofuran

Concen: 217.52 ug/l





RT: 7.567 min Scan# 960

Delta R.T. 0.006 min

Lab File: VD068996.D

Acq: 28 Apr 2021 11:03

Tgt Ion:	42	Resp:	85361
Ion	Ratio	Lower	Upper
42	100		
72	61.3	45.2	67.8
71	56.5	40.8	61.2



#30

Chloroform

Concen: 46.31 ug/l




RT: 7.708 min Scan# 984


Delta R.T. 0.000 min

Lab File: VD068996.D

Acq: 28 Apr 2021 11:03

Tgt Ion:	83	Resp:	294304
Ion	Ratio	Lower	Upper
83	100		
85	64.8	53.3	79.9





#31

Cyclohexane

Concen: 43.01 ug/l

RT: 7.985 min Scan# 1031





Delta R.T. -0.000 min

Lab File: VD068996.D

Acq: 28 Apr 2021 11:03

Tgt Ion: 56 Resp: 213132

Ion	Ratio	Lower	Upper
56	100		
69	37.9	30.0	45.0
84	107.9	82.0	123.0



#32

1,1,1-Trichloroethane

Concen: 48.97 ug/l

RT: 7.903 min Scan# 1017




Delta R.T. 0.001 min


Lab File: VD068996.D

Acq: 28 Apr 2021 11:03

Tgt Ion: 97 Resp: 271969

Ion	Ratio	Lower	Upper
97	100		
99	63.3	50.8	76.2
61	41.1	33.7	50.5









#33
1,2-Dichloroethane-d4
Concen: 44.03 ug/l
RT: 8.332 min Scan# 1090
Delta R.T. -0.000 min
Lab File: VD068996.D
Acq: 28 Apr 2021 11:03




Instrument :
MSVOA_D
ClientSampled :


Tgt Ion: 65 Resp: 131120
Ion Ratio Lower Upper
65 100
67 52.7 0.0 106.0



#34
1,4-Difluorobenzene
Concen: 50.00 ug/l
RT: 8.867 min Scan# 1181
Delta R.T. 0.000 min
Lab File: VD068996.D
Acq: 28 Apr 2021 11:03

Tgt Ion:114 Resp: 436979
Ion Ratio Lower Upper
114 100
63 17.2 0.0 38.2
88 14.6 0.0 29.4





#35

Dibromofluoromethane

Concen: 52.26 ug/l

RT: 7.914 min Scan# 1019

Delta R.T. -0.006 min

Lab File: VD068996.D

Acq: 28 Apr 2021 11:03





Tgt Ion:113 Resp: 148532

Ion Ratio Lower Upper

113 100

111 103.7 82.5 123.7

192 20.2 15.1 22.7



#36

1,1-Dichloropropene

Concen: 50.56 ug/l

RT: 8.114 min Scan# 1053

Delta R.T. 0.000 min

Lab File: VD068996.D

Acq: 28 Apr 2021 11:03




Tgt Ion: 75 Resp: 223616


Ion Ratio Lower Upper

75 100

110 38.7 18.1 54.4

77 31.3 24.7 37.1









#37
Ethyl Acetate
Concen: 52.18 ug/l
RT: 7.291 min Scan# 913
Delta R.T. -0.006 min
Lab File: VD068996.D
Acq: 28 Apr 2021 11:03




Instrument :
MSVOA_D
ClientSampleId :


Tgt Ion: 43 Resp: 67788
Ion Ratio Lower Upper
43 100
61 15.4 14.2 21.4
70 11.8 9.4 14.0



#38
Carbon Tetrachloride
Concen: 55.87 ug/l
RT: 8.097 min Scan# 1050
Delta R.T. -0.000 min
Lab File: VD068996.D
Acq: 28 Apr 2021 11:03

Tgt Ion: 117 Resp: 244652
Ion Ratio Lower Upper
117 100
119 96.3 76.6 115.0
121 29.8 26.1 39.1





#39

Methylcyclohexane

Concen: 53.99 ug/l

RT: 9.355 min Scan# 1264

Delta R.T. 0.000 min





Lab File: VD068996.D

Acq: 28 Apr 2021 11:03

Instrument :
MSVOA_D
ClientSampled :

Tgt Ion: 83 Resp: 278499

Ion	Ratio	Lower	Upper
83	100		
55	62.4	51.7	77.5
98	52.2	38.8	58.2



#40

Benzene

Concen: 49.99 ug/l

RT: 8.350 min Scan# 1093




Delta R.T. -0.000 min


Lab File: VD068996.D

Acq: 28 Apr 2021 11:03

Tgt Ion: 78 Resp: 647083

Ion	Ratio	Lower	Upper
78	100		
77	23.4	18.2	27.2









#41
Methacrylonitrile
Concen: 115.43 ug/l
RT: 7.561 min Scan# 959
Delta R.T. 0.035 min
Lab File: VD068996.D
Acq: 28 Apr 2021 11:03




Instrument :
MSVOA_D
ClientSampleId :


Tgt Ion: 41 Resp: 88780
Ion Ratio Lower Upper
41 100
39 55.1 43.2 64.8
67 0.0 63.9 95.9#
52 0.0 26.2 39.4#



#42
1,2-Dichloroethane
Concen: 49.68 ug/l
RT: 8.420 min Scan# 1105
Delta R.T. -0.006 min
Lab File: VD068996.D
Acq: 28 Apr 2021 11:03

Tgt Ion: 62 Resp: 169979
Ion Ratio Lower Upper
62 100
98 10.7 0.0 22.6









#43

Isopropyl Acetate
Concen: 49.08 ug/l
RT: 8.450 min Scan# 1110
Delta R.T. -0.005 min
Lab File: VD068996.D
Acq: 28 Apr 2021 11:03

Instrument :
MSVOA_D
ClientSampled :




Tgt Ion: 43 Resp: 129985
Ion Ratio Lower Upper
43 100
61 37.5 29.5 44.3
87 17.7 13.5 20.3




#44

Trichloroethene
Concen: 53.34 ug/l
RT: 9.114 min Scan# 1223
Delta R.T. 0.000 min
Lab File: VD068996.D
Acq: 28 Apr 2021 11:03

Tgt Ion:130 Resp: 190674
Ion Ratio Lower Upper
130 100
95 93.5 0.0 203.8









#45

1,2-Dichloropropane
Concen: 49.20 ug/l
RT: 9.385 min Scan# 1269
Delta R.T. -0.006 min
Lab File: VD068996.D
Acq: 28 Apr 2021 11:03

Instrument :
MSVOA_D
ClientSampled :




Tgt Ion: 63 Resp: 152096
Ion Ratio Lower Upper
63 100
65 33.5 27.0 40.4




#46

Dibromomethane
Concen: 50.93 ug/l
RT: 9.473 min Scan# 1284
Delta R.T. 0.000 min
Lab File: VD068996.D
Acq: 28 Apr 2021 11:03

Tgt Ion: 93 Resp: 87703
Ion Ratio Lower Upper
93 100
95 85.4 64.5 96.7
174 99.6 75.3 112.9









#47
Bromodichloromethane
Concen: 50.85 ug/l
RT: 9.661 min Scan# 1316
Delta R.T. -0.006 min
Lab File: VD068996.D
Acq: 28 Apr 2021 11:03




Instrument :
MSVOA_D
ClientSampled :


Tgt Ion: 83 Resp: 228798
Ion Ratio Lower Upper
83 100
85 65.6 51.0 76.6
127 8.9 7.1 10.7



#48
Methyl methacrylate
Concen: 48.67 ug/l
RT: 9.455 min Scan# 1281
Delta R.T. 0.000 min
Lab File: VD068996.D
Acq: 28 Apr 2021 11:03

Tgt Ion: 41 Resp: 63524
Ion Ratio Lower Upper
41 100
69 113.4 83.0 124.6
39 63.6 47.0 70.6





#49

1,4-Dioxane

Concen: 1104.52 ug/l

RT: 9.467 min Scan# 1283





Delta R.T. 0.000 min

Lab File: VD068996.D

Acq: 28 Apr 2021 11:03

Instrument :
MSVOA_D
ClientSampleId :

Tgt Ion:	88	Resp:	20139
Ion	Ratio	Lower	Upper
88	100		
43	27.7	25.8	38.8
58	58.6	54.0	81.0



#50

Toluene-d8

Concen: 51.05 ug/l




RT: 10.338 min Scan# 1431


Delta R.T. -0.000 min

Lab File: VD068996.D

Acq: 28 Apr 2021 11:03

Tgt Ion:	98	Resp:	547375
Ion	Ratio	Lower	Upper
98	100		
100	64.5	51.8	77.6









#51
4-Methyl-2-Pentanone
Concen: 253.74 ug/l
RT: 10.226 min Scan# 1412
Delta R.T. -0.000 min
Lab File: VD068996.D
Acq: 28 Apr 2021 11:03




Instrument :
MSVOA_D
ClientSampled :


Tgt Ion: 43 Resp: 338721
Ion Ratio Lower Upper
43 100
58 45.9 37.3 55.9



#52
Toluene
Concen: 53.69 ug/l
RT: 10.402 min Scan# 1442
Delta R.T. 0.000 min
Lab File: VD068996.D
Acq: 28 Apr 2021 11:03

Tgt Ion: 92 Resp: 440033
Ion Ratio Lower Upper
92 100
91 171.5 139.1 208.7









#53
t-1,3-Dichloropropene
Concen: 51.38 ug/l
RT: 10.620 min Scan# 1479
Delta R.T. 0.000 min
Lab File: VD068996.D
Acq: 28 Apr 2021 11:03




Instrument :
MSVOA_D
ClientSampled :


Tgt Ion: 75 Resp: 204878
Ion Ratio Lower Upper
75 100
77 31.5 25.1 37.7



#54
cis-1,3-Dichloropropene
Concen: 51.13 ug/l
RT: 10.091 min Scan# 1389
Delta R.T. -0.000 min
Lab File: VD068996.D
Acq: 28 Apr 2021 11:03

Tgt Ion: 75 Resp: 243903
Ion Ratio Lower Upper
75 100
77 32.9 24.7 37.1
39 36.7 30.2 45.2









#55
1,1,2-Trichloroethane
Concen: 51.49 ug/l
RT: 10.796 min Scan# 1509
Delta R.T. 0.000 min
Lab File: VD068996.D
Acq: 28 Apr 2021 11:03




Instrument :
MSVOA_D
ClientSampleId :


Tgt Ion:	97	Resp:	125389
Ion	Ratio	Lower	Upper
97	100		
83	83.5	65.7	98.5
85	57.5	44.3	66.5
99	65.5	51.4	77.2



#56
Ethyl methacrylate
Concen: 53.27 ug/l
RT: 10.661 min Scan# 1486
Delta R.T. 0.000 min
Lab File: VD068996.D
Acq: 28 Apr 2021 11:03

Tgt Ion:	69	Resp:	133533
Ion	Ratio	Lower	Upper
69	100		
41	50.7	42.2	63.4
39	30.0	24.2	36.2









#57
1,3-Dichloropropane
Concen: 49.87 ug/l
RT: 10.944 min Scan# 1534
Delta R.T. 0.001 min
Lab File: VD068996.D
Acq: 28 Apr 2021 11:03




Instrument :
MSVOA_D
ClientSampled :


Tgt Ion: 76 Resp: 199962
Ion Ratio Lower Upper
76 100
78 33.2 25.8 38.6



#58
2-Chloroethyl Vinyl ether
Concen: 258.32 ug/l
RT: 9.944 min Scan# 1364
Delta R.T. -0.000 min
Lab File: VD068996.D
Acq: 28 Apr 2021 11:03

Tgt Ion: 63 Resp: 345535
Ion Ratio Lower Upper
63 100
106 32.5 24.6 36.8









#59
2-Hexanone
Concen: 270.47 ug/l
RT: 10.985 min Scan# 1541
Delta R.T. -0.000 min
Lab File: VD068996.D
Acq: 28 Apr 2021 11:03




Instrument :
MSVOA_D
ClientSampled :


Tgt Ion: 43 Resp: 241582
Ion Ratio Lower Upper
43 100
58 64.2 31.3 93.8



#60
Dibromochloromethane
Concen: 54.38 ug/l
RT: 11.138 min Scan# 1567
Delta R.T. -0.000 min
Lab File: VD068996.D
Acq: 28 Apr 2021 11:03

Tgt Ion:129 Resp: 166033
Ion Ratio Lower Upper
129 100
127 77.0 38.8 116.3









#61
1,2-Dibromoethane
Concen: 51.04 ug/l
RT: 11.244 min Scan# 1585
Delta R.T. -0.005 min
Lab File: VD068996.D
Acq: 28 Apr 2021 11:03




Instrument :
MSVOA_D
ClientSampled :


Tgt Ion:107 Resp: 120788
Ion Ratio Lower Upper
107 100
109 94.7 74.9 112.3



#62
4-Bromofluorobenzene
Concen: 50.30 ug/l
RT: 12.632 min Scan# 1821
Delta R.T. 0.006 min
Lab File: VD068996.D
Acq: 28 Apr 2021 11:03

Tgt Ion: 95 Resp: 179132
Ion Ratio Lower Upper
95 100
174 86.8 0.0 168.0
176 83.0 0.0 160.2









#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 11.643 min Scan# 1653
Delta R.T. 0.000 min
Lab File: VD068996.D
Acq: 28 Apr 2021 11:03




Instrument :
MSVOA_D
ClientSampled :


Tgt Ion:117 Resp: 405730
Ion Ratio Lower Upper
117 100
82 54.3 43.5 65.3
119 32.6 26.7 40.1



#64
Tetrachloroethene
Concen: 55.08 ug/l
RT: 10.873 min Scan# 1522
Delta R.T. -0.006 min
Lab File: VD068996.D
Acq: 28 Apr 2021 11:03

Tgt Ion:164 Resp: 156654
Ion Ratio Lower Upper
164 100
166 137.2 103.0 154.6
129 99.9 77.4 116.0
131 94.9 70.4 105.6









#65
Chlorobenzene
Concen: 53.73 ug/l
RT: 11.667 min Scan# 1657
Delta R.T. 0.000 min
Lab File: VD068996.D
Acq: 28 Apr 2021 11:03




Instrument :
MSVOA_D
ClientSampled :


Tgt Ion:112 Resp: 471873
Ion Ratio Lower Upper
112 100
114 31.4 25.2 37.8



#66
1,1,1,2-Tetrachloroethane
Concen: 56.00 ug/l
RT: 11.743 min Scan# 1670
Delta R.T. 0.005 min
Lab File: VD068996.D
Acq: 28 Apr 2021 11:03

Tgt Ion:131 Resp: 176846
Ion Ratio Lower Upper
131 100
133 92.6 47.0 141.2
119 64.4 32.8 98.3









#67
Ethyl Benzene
Concen: 55.18 ug/l
RT: 11.743 min Scan# 1670
Delta R.T. 0.000 min
Lab File: VD068996.D
Acq: 28 Apr 2021 11:03




Instrument :
MSVOA_D
ClientSampled :


Tgt Ion: 91 Resp: 839198
Ion Ratio Lower Upper
91 100
106 32.2 25.1 37.7



#68
m/p-Xylenes
Concen: 110.52 ug/l
RT: 11.849 min Scan# 1688
Delta R.T. 0.000 min
Lab File: VD068996.D
Acq: 28 Apr 2021 11:03

Tgt Ion:106 Resp: 648594
Ion Ratio Lower Upper
106 100
91 196.0 158.6 238.0









#69
o-Xylene
Concen: 54.39 ug/l
RT: 12.179 min Scan# 1744
Delta R.T. -0.000 min
Lab File: VD068996.D
Acq: 28 Apr 2021 11:03




Instrument :
MSVOA_D
ClientSampled :


Tgt Ion:106 Resp: 290228
Ion Ratio Lower Upper
106 100
91 211.3 105.0 315.0



#70
Styrene
Concen: 55.75 ug/l
RT: 12.191 min Scan# 1746
Delta R.T. 0.001 min
Lab File: VD068996.D
Acq: 28 Apr 2021 11:03

Tgt Ion:104 Resp: 518440
Ion Ratio Lower Upper
104 100
78 45.5 36.9 55.3
103 54.4 44.3 66.5









#71
Bromoform
Concen: 54.57 ug/l
RT: 12.355 min Scan# 1774
Delta R.T. 0.000 min
Lab File: VD068996.D
Acq: 28 Apr 2021 11:03




Instrument :
MSVOA_D
ClientSampleId :


Tgt Ion:173 Resp: 92315
Ion Ratio Lower Upper
173 100
175 50.5 23.6 71.0
254 0.1 0.0 0.0#



#72
1,4-Dichlorobenzene-d4
Concen: 50.00 ug/l
RT: 13.573 min Scan# 1981
Delta R.T. -0.000 min
Lab File: VD068996.D
Acq: 28 Apr 2021 11:03

Tgt Ion:152 Resp: 201834
Ion Ratio Lower Upper
152 100
115 59.1 29.3 87.8
150 177.6 0.0 347.6









#73
Isopropylbenzene
Concen: 56.22 ug/l
RT: 12.479 min Scan# 1795
Delta R.T. -0.000 min
Lab File: VD068996.D
Acq: 28 Apr 2021 11:03




Instrument :
MSVOA_D
ClientSampled :


Tgt Ion:105 Resp: 828588
Ion Ratio Lower Upper
105 100
120 26.6 13.4 40.2



#74
N-ethyl acetate
Concen: 48.58 ug/l
RT: 12.285 min Scan# 1762
Delta R.T. -0.000 min
Lab File: VD068996.D
Acq: 28 Apr 2021 11:03

Tgt Ion: 43 Resp: 121408
Ion Ratio Lower Upper
43 100
70 55.1 41.9 62.9
55 29.2 22.8 34.2
61 28.7 22.7 34.1









#75
1,1,2,2-Tetrachloroethane
Concen: 51.00 ug/l
RT: 12.726 min Scan# 1837
Delta R.T. -0.000 min
Lab File: VD068996.D
Acq: 28 Apr 2021 11:03




Instrument :
MSVOA_D
ClientSampled :


Tgt Ion: 83 Resp: 140489
Ion Ratio Lower Upper
83 100
131 9.9 5.1 15.3
85 63.2 32.0 96.0



#76
1,2,3-Trichloropropane
Concen: 93.25 ug/l
RT: 12.779 min Scan# 1846
Delta R.T. -0.000 min
Lab File: VD068996.D
Acq: 28 Apr 2021 11:03

Tgt Ion: 75 Resp: 171943
Ion Ratio Lower Upper
75 100
77 0.0 0.0 0.0









#77
Bromobenzene
Concen: 53.36 ug/l
RT: 12.755 min Scan# 1842
Delta R.T. -0.006 min
Lab File: VD068996.D
Acq: 28 Apr 2021 11:03




Instrument :
MSVOA_D
ClientSampleId :


Tgt Ion:156 Resp: 189034
Ion Ratio Lower Upper
156 100
77 167.2 88.9 266.8
158 97.6 47.7 143.1



#78
n-propylbenzene
Concen: 55.03 ug/l
RT: 12.814 min Scan# 1852
Delta R.T. 0.000 min
Lab File: VD068996.D
Acq: 28 Apr 2021 11:03

Tgt Ion: 91 Resp: 985127
Ion Ratio Lower Upper
91 100
120 23.4 11.4 34.2









#79
2-Chlorotoluene
Concen: 52.18 ug/l
RT: 12.902 min Scan# 1867
Delta R.T. 0.000 min
Lab File: VD068996.D
Acq: 28 Apr 2021 11:03




Instrument :
MSVOA_D
ClientSampled :


Tgt Ion: 91 Resp: 521745
Ion Ratio Lower Upper
91 100
126 35.8 17.1 51.3



#80
1,3,5-Trimethylbenzene
Concen: 55.33 ug/l
RT: 12.955 min Scan# 1876
Delta R.T. 0.000 min
Lab File: VD068996.D
Acq: 28 Apr 2021 11:03

Tgt Ion:105 Resp: 681704
Ion Ratio Lower Upper
105 100
120 50.2 24.9 74.7









#81
trans-1,4-Dichloro-2-butene
Concen: 51.08 ug/l
RT: 12.526 min Scan# 1803
Delta R.T. -0.000 min
Lab File: VD068996.D
Acq: 28 Apr 2021 11:03




Instrument :
MSVOA_D
ClientSampleId :


Tgt Ion: 75 Resp: 40582
Ion Ratio Lower Upper
75 100
53 84.2 66.8 100.2
89 46.5 36.6 54.8



#82
4-Chlorotoluene
Concen: 52.39 ug/l
RT: 13.002 min Scan# 1884
Delta R.T. 0.000 min
Lab File: VD068996.D
Acq: 28 Apr 2021 11:03

Tgt Ion: 91 Resp: 555277
Ion Ratio Lower Upper
91 100
126 35.9 17.2 51.5









#83
tert-Butylbenzene
Concen: 55.40 ug/l
RT: 13.220 min Scan# 1921
Delta R.T. -0.000 min
Lab File: VD068996.D
Acq: 28 Apr 2021 11:03




Instrument :
MSVOA_D
ClientSampleId :


Tgt Ion:119 Resp: 579219
Ion Ratio Lower Upper
119 100
91 62.9 31.6 95.0
134 25.9 13.4 40.2



#84
1,2,4-Trimethylbenzene
Concen: 54.75 ug/l
RT: 13.261 min Scan# 1928
Delta R.T. -0.006 min
Lab File: VD068996.D
Acq: 28 Apr 2021 11:03

Tgt Ion:105 Resp: 681252
Ion Ratio Lower Upper
105 100
120 46.4 22.6 67.7









#85
 sec-Butylbenzene
 Concen: 56.25 ug/l
 RT: 13.396 min Scan# 1951
 Delta R.T. 0.000 min
 Lab File: VD068996.D
 Acq: 28 Apr 2021 11:03




Instrument :
 MSVOA_D
 ClientSampled :


Tgt Ion:105 Resp: 829742
 Ion Ratio Lower Upper
 105 100
 134 20.2 10.0 29.8



#86
 p-Isopropyltoluene
 Concen: 56.70 ug/l
 RT: 13.508 min Scan# 1970
 Delta R.T. -0.006 min
 Lab File: VD068996.D
 Acq: 28 Apr 2021 11:03

Tgt Ion:119 Resp: 763929
 Ion Ratio Lower Upper
 119 100
 134 26.3 13.3 39.9
 91 21.8 11.1 33.1









#87
 1,3-Dichlorobenzene
 Concen: 53.51 ug/l
 RT: 13.514 min Scan# 1971
 Delta R.T. -0.000 min
 Lab File: VD068996.D
 Acq: 28 Apr 2021 11:03




Instrument :
 MSVOA_D
 ClientSampleId :


Tgt Ion:146 Resp: 375602
 Ion Ratio Lower Upper
 146 100
 111 39.2 19.3 57.8
 148 64.3 30.9 92.8



#88
 1,4-Dichlorobenzene
 Concen: 52.86 ug/l
 RT: 13.590 min Scan# 1984
 Delta R.T. 0.000 min
 Lab File: VD068996.D
 Acq: 28 Apr 2021 11:03

Tgt Ion:146 Resp: 367223
 Ion Ratio Lower Upper
 146 100
 111 37.4 19.1 57.1
 148 64.1 31.6 95.0









#89
n-Butylbenzene
Concen: 56.43 ug/l
RT: 13.838 min Scan# 2026
Delta R.T. 0.001 min
Lab File: VD068996.D
Acq: 28 Apr 2021 11:03




Instrument :
MSVOA_D
ClientSampled :


Tgt Ion: 91 Resp: 716178
Ion Ratio Lower Upper
91 100
92 53.0 26.7 80.1
134 25.9 12.9 38.6



#90
Hexachloroethane
Concen: 53.18 ug/l
RT: 14.108 min Scan# 2072
Delta R.T. 0.000 min
Lab File: VD068996.D
Acq: 28 Apr 2021 11:03

Tgt Ion: 117 Resp: 142716
Ion Ratio Lower Upper
117 100
201 75.0 35.4 106.2









#91
1,2-Dichlorobenzene
Concen: 52.53 ug/l
RT: 13.885 min Scan# 2034
Delta R.T. 0.001 min
Lab File: VD068996.D
Acq: 28 Apr 2021 11:03




Instrument :
MSVOA_D
ClientSampled :


Tgt Ion:146 Resp: 319051
Ion Ratio Lower Upper
146 100
111 39.7 20.3 60.9
148 64.0 32.2 96.6



#92
1,2-Dibromo-3-Chloropropane
Concen: 50.97 ug/l
RT: 14.496 min Scan# 2138
Delta R.T. 0.000 min
Lab File: VD068996.D
Acq: 28 Apr 2021 11:03

Tgt Ion: 75 Resp: 21342
Ion Ratio Lower Upper
75 100
155 87.4 46.8 140.3
157 111.4 59.8 179.3









#93
1,2,4-Trichlorobenzene
Concen: 56.46 ug/l
RT: 15.155 min Scan# 2250
Delta R.T. 0.000 min
Lab File: VD068996.D
Acq: 28 Apr 2021 11:03




Instrument :
MSVOA_D
ClientSampled :


Tgt Ion:180 Resp: 212197
Ion Ratio Lower Upper
180 100
182 95.3 47.9 143.8
145 29.8 15.2 45.6



#94
Hexachlorobutadiene
Concen: 59.06 ug/l
RT: 15.261 min Scan# 2268
Delta R.T. -0.000 min
Lab File: VD068996.D
Acq: 28 Apr 2021 11:03

Tgt Ion:225 Resp: 127086
Ion Ratio Lower Upper
225 100
223 61.5 30.6 91.6
227 63.9 31.4 94.0









#95
Naphthalene
Concen: 55.97 ug/l
RT: 15.396 min Scan# 2291
Delta R.T. 0.000 min
Lab File: VD068996.D
Acq: 28 Apr 2021 11:03

Instrument :
MSVOA_D
ClientSampleId :

Tgt Ion:128 Resp: 366431
Ion Ratio Lower Upper
128 100
127 13.4 10.2 15.4
129 10.6 8.9 13.3



#96
1,2,3-Trichlorobenzene
Concen: 55.86 ug/l
RT: 15.590 min Scan# 2324
Delta R.T. 0.000 min
Lab File: VD068996.D
Acq: 28 Apr 2021 11:03

Tgt Ion:180 Resp: 183626
Ion Ratio Lower Upper
180 100
182 95.4 46.5 139.3
145 31.0 15.8 47.4

