

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD042821\
 Data File : VD069003.D
 Acq On : 28 Apr 2021 14:43
 Operator : VA/SY
 Sample : M2199-01
 Misc : 5.06G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SAMPLE-1-GRAB

Quant Time: Apr 29 06:31:59 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D040721S.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 08 03:50:31 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	274514	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	472077	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.644	117	525172	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	216929	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	119665	41.25	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	82.50%
35) Dibromofluoromethane	7.914	113	150377	48.97	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	97.94%
50) Toluene-d8	10.338	98	583645	50.39	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	100.78%
62) 4-Bromofluorobenzene	12.626	95	221856	57.67	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	115.34%
Target Compounds						
16) Acetone	4.162	43	13853	34.48	ug/l	91
17) Carbon Disulfide	4.409	76	130988	14.73	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD042821\
Data File : VD069003.D
Acq On : 28 Apr 2021 14:43
Operator : VA/SY
Sample : M2199-01
Misc : 5.06G/5.00ml/MSVOA_D/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SAMPLE-1-GRAB

Quant Time: Apr 29 06:31:59 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D040721S.M
Quant Title : SW846 8260
QLast Update : Thu Apr 08 03:50:31 2021
Response via : Initial Calibration

