

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD042821\
 Data File : VD069005.D
 Acq On : 28 Apr 2021 15:40
 Operator : VA/SY
 Sample : M2188-01
 Misc : 6.77G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP-9A-8-COMP

Quant Time: Apr 29 06:32:31 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D040721S.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 08 03:50:31 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.985	168	206250	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	366198	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.644	117	461161	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	232746	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	106120	48.68	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery =	97.36%		
35) Dibromofluoromethane	7.914	113	126260	53.01	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery =	106.02%		
50) Toluene-d8	10.338	98	469922	52.30	ug/l	0.00
Spiked Amount 50.000	Range 49 - 140		Recovery =	104.60%		
62) 4-Bromofluorobenzene	12.632	95	217786	72.98	ug/l	0.00
Spiked Amount 50.000	Range 25 - 144		Recovery =	145.96%#		
Target Compounds					Qvalue	
16) Acetone	4.162	43	20262	67.13	ug/l	89
25) 2-Butanone	7.208	43	4266	9.99	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD042821\
Data File : VD069005.D
Acq On : 28 Apr 2021 15:40
Operator : VA/SY
Sample : M2188-01
Misc : 6.77G/5.00ml/MSVOA_D/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP-9A-8-COMP

Quant Time: Apr 29 06:32:31 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D040721S.M
Quant Title : SW846 8260
QLast Update : Thu Apr 08 03:50:31 2021
Response via : Initial Calibration

