

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD042821\
 Data File : VD069011.D
 Acq On : 28 Apr 2021 18:31
 Operator : VA/SY
 Sample : M2213-02
 Misc : 5.12G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BG5B5

Quant Time: Apr 29 06:34:12 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D040721S.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 08 03:50:31 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	246897	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	438208	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.644	117	500849	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	215552	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	111312	42.66	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	85.32%
35) Dibromofluoromethane	7.914	113	136858	48.02	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	96.04%
50) Toluene-d8	10.338	98	539457	50.17	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	100.34%
62) 4-Bromofluorobenzene	12.626	95	215321	60.30	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	120.60%
Target Compounds						
16) Acetone	4.162	43	9898	27.39	ug/l	92
17) Carbon Disulfide	4.409	76	18491	2.31	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD042821\
Data File : VD069011.D
Acq On : 28 Apr 2021 18:31
Operator : VA/SY
Sample : M2213-02
Misc : 5.12G/5.00ml/MSVOA_D/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BG5B5

Quant Time: Apr 29 06:34:12 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D040721S.M
Quant Title : SW846 8260
QLast Update : Thu Apr 08 03:50:31 2021
Response via : Initial Calibration

