

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD042821\
 Data File : VD069013.D
 Acq On : 28 Apr 2021 19:28
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 29 06:34:41 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D040721S.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 08 03:50:31 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	86	0.00
2 T	Dichlorodifluoromethane	0.495	0.415	16.2	74	0.00
3 P	Chloromethane	0.582	0.500	14.1	76	0.00
4 C	Vinyl Chloride	0.732	0.698	4.6#	84	0.00
5 T	Bromomethane	0.547	0.561	-2.6	96	-0.01
6 T	Chloroethane	0.481	0.524	-8.9	96	0.00
7 T	Trichlorofluoromethane	0.861	0.873	-1.4	92	0.00
8 T	Diethyl Ether	0.233	0.248	-6.4	85	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.520	0.525	-1.0	84	0.00
10 T	Methyl Iodide	0.538	0.565	-5.0	77	0.00
11 T	Tert butyl alcohol	0.034	0.032	5.9	92	0.00
12 CM	1,1-Dichloroethene	0.501	0.490	2.2#	82	-0.01
13 T	Acrolein	0.021	0.015	28.6#	60	0.00
14 T	Allyl chloride	0.564	0.508	9.9	75	0.00
15 T	Acrylonitrile	0.104	0.101	2.9	82	0.00
16 T	Acetone	0.073	0.066	9.6	78	0.00
17 T	Carbon Disulfide	1.620	1.405	13.3	71	0.00
18 T	Methyl Acetate	0.201	0.211	-5.0	90	0.00
19 T	Methyl tert-butyl Ether	1.110	1.177	-6.0	88	0.00
20 T	Methylene Chloride	0.679	0.696	-2.5	94	-0.01
21 T	trans-1,2-Dichloroethene	0.643	0.632	1.7	85	0.00
22 T	Diisopropyl ether	1.367	1.307	4.4	79	0.00
23 T	Vinyl Acetate	0.724	0.719	0.7	79	0.00
24 P	1,1-Dichloroethane	1.030	1.013	1.7	85	0.00
25 T	2-Butanone	0.104	0.107	-2.9	85	0.00
26 T	2,2-Dichloropropane	0.842	0.845	-0.4	90	0.00
27 T	cis-1,2-Dichloroethene	0.639	0.729	-14.1	97	0.00
28 T	Bromochloromethane	0.372	0.353	5.1	81	0.00
29 T	Tetrahydrofuran	0.070	0.069	1.4	80	0.00
30 C	Chloroform	1.128	1.161	-2.9#	89	0.00
31 T	Cyclohexane	0.879	0.709	19.3	71	0.00
32 T	1,1,1-Trichloroethane	0.986	0.968	1.8	85	0.00
33 S	1,2-Dichloroethane-d4	0.528	0.522	1.1	87	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	84	0.00
35 S	Dibromofluoromethane	0.325	0.361	-11.1	94	0.00
36 T	1,1-Dichloropropene	0.506	0.489	3.4	81	0.00
37 T	Ethyl Acetate	0.149	0.161	-8.1	95	0.00
38 T	Carbon Tetrachloride	0.501	0.520	-3.8	86	0.00
39 T	Methylcyclohexane	0.590	0.564	4.4	76	0.00
40 TM	Benzene	1.481	1.494	-0.9	85	0.00
41 T	Methacrylonitrile	0.088	0.090	-2.3	104	0.01
42 TM	1,2-Dichloroethane	0.391	0.402	-2.8	86	0.00
43 T	Isopropyl Acetate	0.303	0.311	-2.6	85	0.00
44 TM	Trichloroethene	0.409	0.427	-4.4	88	0.00
45 C	1,2-Dichloropropane	0.354	0.348	1.7#	83	0.00
46 T	Dibromomethane	0.197	0.215	-9.1	90	0.00
47 T	Bromodichloromethane	0.515	0.550	-6.8	91	0.00
48 T	Methyl methacrylate	0.149	0.142	4.7	74	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD042821\
 Data File : VD069013.D
 Acq On : 28 Apr 2021 19:28
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 29 06:34:41 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D040721S.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 08 03:50:31 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	93	0.00
50 S	Toluene-d8	1.227	1.292	-5.3	88	0.00
51 T	4-Methyl-2-Pentanone	0.153	0.163	-6.5	84	0.00
52 CM	Toluene	0.938	1.007	-7.4#	87	0.00
53 T	t-1,3-Dichloropropene	0.456	0.474	-3.9	85	0.00
54 T	cis-1,3-Dichloropropene	0.546	0.563	-3.1	85	0.00
55 T	1,1,2-Trichloroethane	0.279	0.303	-8.6	91	0.00
56 T	Ethyl methacrylate	0.287	0.323	-12.5	87	0.00
57 T	1,3-Dichloropropane	0.459	0.498	-8.5	90	0.00
58 T	2-Chloroethyl Vinyl ether	0.140	0.159	-13.6	86	0.00
59 T	2-Hexanone	0.102	0.110	-7.8	83	0.00
60 T	Dibromochloromethane	0.349	0.414	-18.6	100	0.00
61 T	1,2-Dibromoethane	0.271	0.301	-11.1	93	0.00
62 S	4-Bromofluorobenzene	0.407	0.440	-8.1	91	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	87	0.00
64 T	Tetrachloroethene	0.350	0.370	-5.7	92	0.00
65 PM	Chlorobenzene	1.082	1.149	-6.2	91	0.00
66 T	1,1,1,2-Tetrachloroethane	0.389	0.447	-14.9	98	0.00
67 C	Ethyl Benzene	1.874	1.952	-4.2#	86	0.00
68 T	m/p-Xylenes	0.723	0.773	-6.9	89	0.00
69 T	o-Xylene	0.658	0.716	-8.8	89	0.00
70 T	Styrene	1.146	1.272	-11.0	90	0.00
71 P	Bromoform	0.208	0.239	-14.9	98	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	92	0.00
73 T	Isopropylbenzene	3.651	3.634	0.5	89	0.00
74 T	N-amyl acetate	0.619	0.575	7.1	82	0.00
75 P	1,1,2,2-Tetrachloroethane	0.682	0.675	1.0	92	0.00
76 T	1,2,3-Trichloropropane	0.457	0.490	-7.2	101	0.00
77 T	Bromobenzene	0.878	0.912	-3.9	98	0.00
78 T	n-propylbenzene	4.434	4.299	3.0	87	0.00
79 T	2-Chlorotoluene	2.477	2.441	1.5	90	0.00
80 T	1,3,5-Trimethylbenzene	3.052	3.104	-1.7	89	0.00
81 T	trans-1,4-Dichloro-2-butene	0.197	0.191	3.0	84	0.00
82 T	4-Chlorotoluene	2.626	2.591	1.3	91	0.00
83 T	tert-Butylbenzene	2.590	2.648	-2.2	91	0.00
84 T	1,2,4-Trimethylbenzene	3.083	3.145	-2.0	90	0.00
85 T	sec-Butylbenzene	3.654	3.686	-0.9	89	0.00
86 T	p-Isopropyltoluene	3.338	3.374	-1.1	89	0.00
87 T	1,3-Dichlorobenzene	1.739	1.780	-2.4	94	0.00
88 T	1,4-Dichlorobenzene	1.721	1.759	-2.2	95	0.00
89 T	n-Butylbenzene	3.144	3.095	1.6	86	0.00
90 T	Hexachloroethane	0.665	0.685	-3.0	95	0.00
91 T	1,2-Dichlorobenzene	1.505	1.530	-1.7	93	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.104	0.097	6.7	88	0.00
93 T	1,2,4-Trichlorobenzene	0.931	0.974	-4.6	96	0.00
94 T	Hexachlorobutadiene	0.533	0.567	-6.4	94	0.00
95 T	Naphthalene	1.622	1.770	-9.1	94	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD042821\
Data File : VD069013.D
Acq On : 28 Apr 2021 19:28
Operator : VA/SY
Sample : VSTDCCC050
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_D
LabSampleId :
VSTDCCC050

Quant Time: Apr 29 06:34:41 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D040721S.M
Quant Title : SW846 8260
QLast Update : Thu Apr 08 03:50:31 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.814	0.871	-7.0	96	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6