

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD042821\
 Data File : VD068997.D
 Acq On : 28 Apr 2021 11:37
 Operator : VA/SY
 Sample : VD0428SBL01
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD0428SBL01

Quant Time: Apr 29 06:29:37 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D040721S.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 08 03:50:31 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	316795	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	543168	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.643	117	646592	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	341696	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	133430	39.85	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	79.70%
35) Dibromofluoromethane	7.920	113	166678	47.18	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	94.36%
50) Toluene-d8	10.338	98	694814	52.13	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	104.26%
62) 4-Bromofluorobenzene	12.632	95	291228	65.79	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	131.58%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD042821\
Data File : VD068997.D
Acq On : 28 Apr 2021 11:37
Operator : VA/SY
Sample : VD0428SBL01
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD0428SBL01

Quant Time: Apr 29 06:29:37 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D040721S.M
Quant Title : SW846 8260
QLast Update : Thu Apr 08 03:50:31 2021
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.00 ug/l
RT: 7.979 min Scan# 1030
Delta R.T. -0.000 min
Lab File: VD068997.D
Acq: 28 Apr 2021 11:37




Instrument :
MSVOA_D
ClientSampleId :
VD0428SBL01


Tgt Ion:168 Resp: 316795
Ion Ratio Lower Upper
168 100
99 53.0 46.2 69.2



#33
1,2-Dichloroethane-d4
Concen: 39.85 ug/l
RT: 8.332 min Scan# 1090
Delta R.T. -0.000 min
Lab File: VD068997.D
Acq: 28 Apr 2021 11:37

Tgt Ion: 65 Resp: 133430
Ion Ratio Lower Upper
65 100
67 53.1 0.0 106.0









#34
1,4-Difluorobenzene
Concen: 50.00 ug/l
RT: 8.867 min Scan# 1181
Delta R.T. 0.000 min
Lab File: VD068997.D
Acq: 28 Apr 2021 11:37




Instrument :
MSVOA_D
ClientSampleId :
VD0428SBL01


Tgt Ion:114 Resp: 543168
Ion Ratio Lower Upper
114 100
63 16.7 0.0 38.2
88 13.8 0.0 29.4



#35
Dibromofluoromethane
Concen: 47.18 ug/l
RT: 7.920 min Scan# 1020
Delta R.T. 0.000 min
Lab File: VD068997.D
Acq: 28 Apr 2021 11:37

Tgt Ion:113 Resp: 166678
Ion Ratio Lower Upper
113 100
111 101.9 82.5 123.7
192 19.8 15.1 22.7









#50
Toluene-d8
Concen: 52.13 ug/l
RT: 10.338 min Scan# 1431
Delta R.T. -0.000 min
Lab File: VD068997.D
Acq: 28 Apr 2021 11:37




Instrument :
MSVOA_D
ClientSampleId :
VD0428SBL01


Tgt Ion: 98 Resp: 694814
Ion Ratio Lower Upper
98 100
100 64.6 51.8 77.6



#62
4-Bromofluorobenzene
Concen: 65.79 ug/l
RT: 12.632 min Scan# 1821
Delta R.T. 0.006 min
Lab File: VD068997.D
Acq: 28 Apr 2021 11:37

Tgt Ion: 95 Resp: 291228
Ion Ratio Lower Upper
95 100
174 88.7 0.0 168.0
176 85.9 0.0 160.2








#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 11.643 min Scan# 1653
Delta R.T. 0.000 min
Lab File: VD068997.D
Acq: 28 Apr 2021 11:37

Instrument :
MSVOA_D
ClientSampleId :
VD0428SBL01

Tgt Ion:117 Resp: 646592
Ion Ratio Lower Upper
117 100
82 50.1 43.5 65.3
119 31.2 26.7 40.1



#72
1,4-Dichlorobenzene-d4
Concen: 50.00 ug/l
RT: 13.573 min Scan# 1981
Delta R.T. -0.000 min
Lab File: VD068997.D
Acq: 28 Apr 2021 11:37

Tgt Ion:152 Resp: 341696
Ion Ratio Lower Upper
152 100
115 56.7 29.3 87.8
150 158.8 0.0 347.6

