

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD042823\
 Data File : VD075884.D
 Acq On : 28 Apr 2023 13:24
 Operator : KP/SY
 Sample : 02525-06
 Misc : 5.20G/10.00ml/MSVOA_D/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 DCCC6

Quant Time: Apr 29 02:58:13 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\SFAMDLM042023SMA.M
 Quant Title : SFAM01.0
 QLast Update : Sat Apr 29 02:40:46 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.775	114	602337	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.581	117	597288	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.522	152	284192	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.275	65	209702	15.870	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	63.480%
7) Chloroethane-d5	2.799	69	202678	17.752	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	71.000%
11) 1,1-Dichloroethene-d2	3.916	65	51114	16.779	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	67.120%
21) 2-Butanone-d5	6.987	46	54326	47.359	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	94.720%
24) Chloroform-d	7.569	84	301508	20.017	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	80.080%
26) 1,2-Dichloroethane-d4	8.228	65	131185	20.395	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	81.560%
32) Benzene-d6	8.199	84	590440	16.961	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	67.840%
36) 1,2-Dichloropropane-d6	9.210	67	183432	19.037	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	76.160%
41) Toluene-d8	10.269	98	545952	16.751	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	67.000%
43) trans-1,3-Dichloroprop...	10.528	79	72702	18.254	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	73.000%
47) 2-Hexanone-d5	10.881	63	48546	41.842	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	83.680%
56) 1,1,2,2-Tetrachloroeth...	12.651	84	160801	21.808	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	87.240%
66) 1,2-Dichlorobenzene-d4	13.816	152	199715	19.868	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	79.480%
Target Compounds					Qvalue	
13) Acetone	4.028	43	21659	22.063	ug/L	99
16) Methylene chloride	4.799	84	17782	1.455	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD042823\
Data File : VD075884.D
Acq On : 28 Apr 2023 13:24
Operator : KP/SY
Sample : 02525-06
Misc : 5.20G/10.00ml/MSVOA_D/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
DCCC6

Quant Time: Apr 29 02:58:13 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\SFAMDLM042023SMA.M
Quant Title : SFAM01.0
QLast Update : Sat Apr 29 02:40:46 2023
Response via : Initial Calibration

