

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD042823\
 Data File : VD075893.D
 Acq On : 28 Apr 2023 18:01
 Operator : KP/SY
 Sample : 02525-13
 Misc : 5.95G/10.00ml/MSVOA_D/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 DCCL0

Quant Time: Apr 29 05:13:42 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\SFAMDLM042023SMA.M
 Quant Title : SFAM01.0
 QLast Update : Sat Apr 29 05:12:16 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	8.775	114	456557	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.581	117	481654	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.522	152	222291	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.276	65	154513	15.427	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	61.720%
7) Chloroethane-d5	2.799	69	156500	18.085	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	72.320%
11) 1,1-Dichloroethene-d2	3.917	65	39315	17.026	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	68.120%
21) 2-Butanone-d5	6.987	46	60464	69.540	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	139.080%#
24) Chloroform-d	7.563	84	257748	22.576	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	90.320%
26) 1,2-Dichloroethane-d4	8.228	65	125133	25.665	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	102.680%
32) Benzene-d6	8.199	84	498934	17.773	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	71.080%
36) 1,2-Dichloropropane-d6	9.210	67	165833	21.342	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	85.360%
41) Toluene-d8	10.269	98	449279	17.095	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	68.360%
43) trans-1,3-Dichloroprop...	10.522	79	67456	21.003	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	84.000%
47) 2-Hexanone-d5	10.875	63	50902	54.406	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	108.820%
56) 1,1,2,2-Tetrachloroeth...	12.651	84	159965	26.903	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	107.600%
66) 1,2-Dichlorobenzene-d4	13.816	152	174804	22.232	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	88.920%
Target Compounds						
13) Acetone	4.028	43	7750	10.416	ug/L	Qvalue 77


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD042823\
Data File : VD075893.D
Acq On : 28 Apr 2023 18:01
Operator : KP/SY
Sample : 02525-13
Misc : 5.95G/10.00ml/MSVOA_D/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
DCCL0


Quant Time: Apr 29 05:13:42 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\SFAMDLM042023SMA.M
Quant Title : SFAM01.0
QLast Update : Sat Apr 29 05:12:16 2023
Response via : Initial Calibration





#13
 Acetone
 Concen: 10.416 ug/L
 RT: 4.028 min Scan# 41
 Delta R.T. 0.012 min
 Lab File: VD075893.D
 Acq: 28 Apr 2023 18:01

Instrument :
 MSVOA_D
 ClientSampleId :
 DCCL0



Tgt Ion: 43 Resp: 7750
 Ion Ratio Lower Upper
 43 100
 58 28.6 0.0 87.4

