

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD042921\
 Data File : VD069034.D
 Acq On : 29 Apr 2021 22:41
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 30 02:59:12 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D040721S.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 08 03:50:31 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	88	0.00
2 T	Dichlorodifluoromethane	0.495	0.414	16.4	76	0.00
3 P	Chloromethane	0.582	0.506	13.1	80	0.00
4 C	Vinyl Chloride	0.732	0.687	6.1#	86	0.00
5 T	Bromomethane	0.547	0.587	-7.3	104	0.00
6 T	Chloroethane	0.481	0.541	-12.5	103	0.00
7 T	Trichlorofluoromethane	0.861	0.893	-3.7	97	0.00
8 T	Diethyl Ether	0.233	0.267	-14.6	95	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.520	0.535	-2.9	88	0.00
10 T	Methyl Iodide	0.538	0.552	-2.6	78	0.00
11 T	Tert butyl alcohol	0.034	0.032	5.9	96	0.00
12 CM	1,1-Dichloroethene	0.501	0.504	-0.6#	87	0.00
13 T	Acrolein	0.021	0.014	33.3#	60	0.00
14 T	Allyl chloride	0.564	0.521	7.6	79	0.00
15 T	Acrylonitrile	0.104	0.110	-5.8	92	0.00
16 T	Acetone	0.073	0.072	1.4	87	0.00
17 T	Carbon Disulfide	1.620	1.425	12.0	74	0.00
18 T	Methyl Acetate	0.201	0.216	-7.5	95	0.00
19 T	Methyl tert-butyl Ether	1.110	1.227	-10.5	95	0.00
20 T	Methylene Chloride	0.679	0.767	-13.0	106	0.00
21 T	trans-1,2-Dichloroethene	0.643	0.635	1.2	88	0.00
22 T	Diisopropyl ether	1.367	1.312	4.0	82	0.00
23 T	Vinyl Acetate	0.724	0.731	-1.0	83	0.00
24 P	1,1-Dichloroethane	1.030	1.000	2.9	87	0.00
25 T	2-Butanone	0.104	0.111	-6.7	90	0.00
26 T	2,2-Dichloropropane	0.842	0.859	-2.0	94	0.00
27 T	cis-1,2-Dichloroethene	0.639	0.729	-14.1	100	0.00
28 T	Bromochloromethane	0.372	0.347	6.7	82	0.00
29 T	Tetrahydrofuran	0.070	0.072	-2.9	86	0.00
30 C	Chloroform	1.128	1.155	-2.4#	92	0.00
31 T	Cyclohexane	0.879	0.700	20.4	73	0.00
32 T	1,1,1-Trichloroethane	0.986	1.011	-2.5	91	0.00
33 S	1,2-Dichloroethane-d4	0.528	0.505	4.4	86	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	84	0.00
35 S	Dibromofluoromethane	0.325	0.363	-11.7	95	0.00
36 T	1,1-Dichloropropene	0.506	0.492	2.8	82	0.00
37 T	Ethyl Acetate	0.149	0.174	-16.8	103	0.00
38 T	Carbon Tetrachloride	0.501	0.537	-7.2	90	0.00
39 T	Methylcyclohexane	0.590	0.570	3.4	77	0.00
40 TM	Benzene	1.481	1.525	-3.0	87	0.00
41 T	Methacrylonitrile	0.088	0.108	-22.7	125	0.00
42 TM	1,2-Dichloroethane	0.391	0.418	-6.9	90	0.00
43 T	Isopropyl Acetate	0.303	0.323	-6.6	89	0.00
44 TM	Trichloroethene	0.409	0.448	-9.5	93	0.00
45 C	1,2-Dichloropropane	0.354	0.367	-3.7#	88	0.00
46 T	Dibromomethane	0.197	0.235	-19.3	99	0.00
47 T	Bromodichloromethane	0.515	0.577	-12.0	96	0.00
48 T	Methyl methacrylate	0.149	0.150	-0.7	78	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD042921\
 Data File : VD069034.D
 Acq On : 29 Apr 2021 22:41
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 30 02:59:12 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D040721S.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 08 03:50:31 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.003	-50.0#	105	0.00
50 S	Toluene-d8	1.227	1.281	-4.4	88	0.00
51 T	4-Methyl-2-Pentanone	0.153	0.175	-14.4	90	0.00
52 CM	Toluene	0.938	1.028	-9.6#	89	0.00
53 T	t-1,3-Dichloropropene	0.456	0.505	-10.7	91	0.00
54 T	cis-1,3-Dichloropropene	0.546	0.593	-8.6	90	0.00
55 T	1,1,2-Trichloroethane	0.279	0.332	-19.0	100	0.00
56 T	Ethyl methacrylate	0.287	0.345	-20.2	93	0.00
57 T	1,3-Dichloropropane	0.459	0.517	-12.6	94	0.00
58 T	2-Chloroethyl Vinyl ether	0.140	0.164	-17.1	89	0.00
59 T	2-Hexanone	0.102	0.118	-15.7	90	0.00
60 T	Dibromochloromethane	0.349	0.427	-22.3	103	0.00
61 T	1,2-Dibromoethane	0.271	0.316	-16.6	99	0.00
62 S	4-Bromofluorobenzene	0.407	0.435	-6.9	90	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	88	0.00
64 T	Tetrachloroethene	0.350	0.384	-9.7	97	0.00
65 PM	Chlorobenzene	1.082	1.186	-9.6	95	0.00
66 T	1,1,1,2-Tetrachloroethane	0.389	0.455	-17.0	102	0.00
67 C	Ethyl Benzene	1.874	2.018	-7.7#	91	0.00
68 T	m/p-Xylenes	0.723	0.794	-9.8	93	0.00
69 T	o-Xylene	0.658	0.749	-13.8	95	0.00
70 T	Styrene	1.146	1.295	-13.0	93	0.00
71 P	Bromoform	0.208	0.249	-19.7	104	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	91	0.00
73 T	Isopropylbenzene	3.651	3.816	-4.5	92	0.00
74 T	N-amyl acetate	0.619	0.630	-1.8	88	0.00
75 P	1,1,2,2-Tetrachloroethane	0.682	0.725	-6.3	97	0.00
76 T	1,2,3-Trichloropropane	0.457	0.556	-21.7	113	0.00
77 T	Bromobenzene	0.878	0.970	-10.5	102	0.00
78 T	n-propylbenzene	4.434	4.541	-2.4	90	0.00
79 T	2-Chlorotoluene	2.477	2.536	-2.4	92	0.00
80 T	1,3,5-Trimethylbenzene	3.052	3.226	-5.7	91	0.00
81 T	trans-1,4-Dichloro-2-butene	0.197	0.205	-4.1	89	0.00
82 T	4-Chlorotoluene	2.626	2.670	-1.7	92	0.00
83 T	tert-Butylbenzene	2.590	2.712	-4.7	92	0.00
84 T	1,2,4-Trimethylbenzene	3.083	3.293	-6.8	92	0.00
85 T	sec-Butylbenzene	3.654	3.813	-4.4	90	0.00
86 T	p-Isopropyltoluene	3.338	3.577	-7.2	93	0.00
87 T	1,3-Dichlorobenzene	1.739	1.884	-8.3	97	0.00
88 T	1,4-Dichlorobenzene	1.721	1.862	-8.2	99	0.00
89 T	n-Butylbenzene	3.144	3.164	-0.6	87	0.00
90 T	Hexachloroethane	0.665	0.692	-4.1	94	0.00
91 T	1,2-Dichlorobenzene	1.505	1.676	-11.4	100	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.104	0.109	-4.8	97	0.00
93 T	1,2,4-Trichlorobenzene	0.931	1.045	-12.2	101	0.00
94 T	Hexachlorobutadiene	0.533	0.599	-12.4	98	0.00
95 T	Naphthalene	1.622	1.958	-20.7	102	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD042921\
Data File : VD069034.D
Acq On : 29 Apr 2021 22:41
Operator : VA/SY
Sample : VSTDCCC050
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_D
LabSampleId :
VSTDCCC050

Quant Time: Apr 30 02:59:12 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D040721S.M
Quant Title : SW846 8260
QLast Update : Thu Apr 08 03:50:31 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.814	0.959	-17.8	104	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6