


Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD042922\
 Data File : VD072731.D
 Acq On : 29 Apr 2022 08:50
 Operator : VA/SY
 Sample : BFB
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D042722S.M
 Title : SW846 8260
 Last Update : Wed Apr 27 16:01:06 2022



AutoFind: Scans 1818, 1819, 1820; Background Corrected with Scan 1806

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.8	49739	PASS
75	95	30	60	52.2	124885	PASS
95	95	100	100	100.0	239275	PASS
96	95	5	9	6.8	16299	PASS
173	174	0.00	2	1.4	2303	PASS
174	95	50	100	71.2	170325	PASS
175	174	5	9	7.6	13023	PASS
176	174	95	101	95.1	162005	PASS
177	176	5	9	7.1	11502	PASS