

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD042923\
 Data File : VD075919.D
 Acq On : 29 Apr 2023 16:21
 Operator : KP/SY
 Sample : 02525-09RE
 Misc : 5.92G/10.00ml/MSVOA_D/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 DCCK6RE

Quant Time: May 01 00:17:09 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\SFAMDLM042023SMA.M
 Quant Title : SFAM01.0
 QLast Update : Mon May 01 00:14:49 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.775	114	467492	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.581	117	493500	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.522	152	219547	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.270	65	179708	17.523	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	70.080%
7) Chloroethane-d5	2.799	69	175728	19.832	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	79.320%
11) 1,1-Dichloroethene-d2	3.917	65	43633	18.454	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	73.800%
21) 2-Butanone-d5	6.981	46	57401	64.473	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	128.940%
24) Chloroform-d	7.569	84	275809	23.593	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	94.360%
26) 1,2-Dichloroethane-d4	8.228	65	124946	25.028	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	100.120%
32) Benzene-d6	8.199	84	515640	17.927	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	71.720%
36) 1,2-Dichloropropane-d6	9.210	67	171069	21.487	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	85.960%
41) Toluene-d8	10.269	98	470562	17.475	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	69.880%
43) trans-1,3-Dichloroprop...	10.528	79	63238	19.217	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	76.880%
47) 2-Hexanone-d5	10.881	63	48882	50.993	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	101.980%
56) 1,1,2,2-Tetrachloroeth...	12.651	84	157119	25.790	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	103.160%
66) 1,2-Dichlorobenzene-d4	13.816	152	178888	23.036	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	92.160%
Target Compounds					Qvalue	
13) Acetone	4.023	43	40967	53.769	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD042923\
Data File : VD075919.D
Acq On : 29 Apr 2023 16:21
Operator : KP/SY
Sample : 02525-09RE
Misc : 5.92G/10.00ml/MSVOA_D/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
DCCK6RE

Quant Time: May 01 00:17:09 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\SFAMDLM042023SMA.M
Quant Title : SFAM01.0
QLast Update : Mon May 01 00:14:49 2023
Response via : Initial Calibration

