


Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD043019\
 Data File : VD062081.D
 Acq On : 30 Apr 2019 9:44
 Operator : FY/SY
 Sample : BFB
 Misc : 5.00qm/5mL/MSVOA D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D042319S.M
 Title : SW846 8260
 Last Update : Wed May 01 08:37:17 2019



AutoFind: Scans 1231, 1232, 1233; Background Corrected with Scan 1226

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.1	17312	PASS
75	95	30	60	40.2	40722	PASS
95	95	100	100	100.0	101173	PASS
96	95	5	9	6.4	6499	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	87.1	88106	PASS
175	174	5	9	6.7	5891	PASS
176	174	95	101	99.4	87557	PASS
177	176	5	9	6.0	5254	PASS