

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD050219\
 Data File : VD062116.D
 Acq On : 2 May 2019 17:34
 Operator : FY/SY
 Sample : K2668-03
 Misc : 4.41g/5mL/MSVOA D/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP-2

Quant Time: May 03 01:49:23 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D042319S.M
 Quant Title : SW846 8260
 QLast Update : Thu May 02 07:55:52 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.02	168	420575	50.00	ug/l	-0.03
34) 1,4-Difluorobenzene	8.19	114	716580	50.00	ug/l	-0.03
63) Chlorobenzene-d5	12.36	117	622532	50.00	ug/l	-0.01
72) 1,4-Dichlorobenzene-d4	14.51	152	269481	50.00	ug/l	-0.01
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.43	65	183794	55.84	ug/l	-0.03
Spiked Amount 50.000			Recovery	=	111.68%	
35) Dibromofluoromethane	6.90	113	247809	50.00	ug/l	-0.03
Spiked Amount 50.000			Recovery	=	100.00%	
50) Toluene-d8	10.39	98	584873	49.60	ug/l	-0.03
Spiked Amount 50.000			Recovery	=	99.20%	
62) 4-Bromofluorobenzene	13.55	95	235676	49.83	ug/l	-0.01
Spiked Amount 50.000			Recovery	=	99.66%	
Target Compounds						
16) Acetone	3.20	43	68451	87.093	ug/l	92
20) Methylene Chloride	3.79	84	34555	6.240	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD050219\
Data File : VD062116.D
Acq On : 2 May 2019 17:34
Operator : FY/SY
Sample : K2668-03
Misc : 4.41g/5mL/MSVOA D/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP-2

Quant Time: May 03 01:49:23 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D042319S.M
Quant Title : SW846 8260
QLast Update : Thu May 02 07:55:52 2019
Response via : Initial Calibration

