

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD050222\
 Data File : VD072757.D
 Acq On : 02 May 2022 20:29
 Operator : VA/SY
 Sample : N2609-10
 Misc : 5.65G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 WC-4

Quant Time: May 03 08:37:27 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D050222S.M
 Quant Title : SW846 8260
 QLast Update : Tue May 03 08:33:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	21033	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	38695	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	41424	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	19028	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	21551	78.934	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	157.860%
35) Dibromofluoromethane	7.908	113	17531	64.509	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	129.020%
50) Toluene-d8	10.332	98	45589	46.766	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	93.540%
62) 4-Bromofluorobenzene	12.620	95	16739	46.051	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	92.100%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD050222\
Data File : VD072757.D
Acq On : 02 May 2022 20:29
Operator : VA/SY
Sample : N2609-10
Misc : 5.65G/5.00ml/MSVOA_D/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
WC-4

Quant Time: May 03 08:37:27 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D050222S.M
Quant Title : SW846 8260
QLast Update : Tue May 03 08:33:49 2022
Response via : Initial Calibration

