

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD050223\
 Data File : VD075956.D
 Acq On : 02 May 2023 11:10
 Operator : KP/SY
 Sample : 02564-11
 Misc : 6.11G/10.00ml/MSVOA_D/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 DCCH1

Quant Time: May 03 02:47:28 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\SFAMDLM042023SMA.M
 Quant Title : SFAM01.0
 QLast Update : Wed May 03 02:37:35 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.781	114	477525	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.587	117	451513	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.522	152	183961	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.275	65	147257	14.057	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	56.240%		
7) Chloroethane-d5	2.799	69	146016	16.132	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	64.520%		
11) 1,1-Dichloroethene-d2	3.922	65	36282	15.023	ug/L	0.00
Spiked Amount 25.000	Range 45 - 110		Recovery =	60.080%		
21) 2-Butanone-d5	6.993	46	48598	53.439	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	106.880%		
24) Chloroform-d	7.569	84	233459	19.551	ug/L	0.00
Spiked Amount 25.000	Range 40 - 150		Recovery =	78.200%		
26) 1,2-Dichloroethane-d4	8.228	65	106670	20.918	ug/L	0.00
Spiked Amount 25.000	Range 70 - 130		Recovery =	83.680%		
32) Benzene-d6	8.204	84	433703	16.480	ug/L	0.00
Spiked Amount 25.000	Range 20 - 135		Recovery =	65.920%		
36) 1,2-Dichloropropane-d6	9.210	67	143571	19.710	ug/L	0.00
Spiked Amount 25.000	Range 70 - 120		Recovery =	78.840%		
41) Toluene-d8	10.269	98	386203	15.676	ug/L	0.00
Spiked Amount 25.000	Range 30 - 130		Recovery =	62.720%		
43) trans-1,3-Dichloroprop...	10.528	79	55370	18.391	ug/L	0.00
Spiked Amount 25.000	Range 30 - 135		Recovery =	73.560%		
47) 2-Hexanone-d5	10.881	63	39737	45.308	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	90.620%		
56) 1,1,2,2-Tetrachloroeth...	12.651	84	131568	23.604	ug/L	0.00
Spiked Amount 25.000	Range 45 - 120		Recovery =	94.400%		
66) 1,2-Dichlorobenzene-d4	13.816	152	153096	23.528	ug/L	0.00
Spiked Amount 25.000	Range 75 - 120		Recovery =	94.120%		
Target Compounds					Qvalue	
13) Acetone	4.022	43	30406	39.070	ug/L	74
16) Methylene chloride	4.805	84	7587	0.783	ug/L	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD050223\
Data File : VD075956.D
Acq On : 02 May 2023 11:10
Operator : KP/SY
Sample : 02564-11
Misc : 6.11G/10.00ml/MSVOA_D/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
DCCH1

Quant Time: May 03 02:47:28 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\SFAMDLM042023SMA.M
Quant Title : SFAM01.0
QLast Update : Wed May 03 02:37:35 2023
Response via : Initial Calibration

