

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD050321\
 Data File : VD069049.D
 Acq On : 03 May 2021 18:43
 Operator : VA/SY
 Sample : M2235-01
 Misc : 4.54G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 OR-02-043021

Quant Time: May 04 04:15:18 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D050321S.M
 Quant Title : SW846 8260
 QLast Update : Tue May 04 03:51:01 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	7621	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	14933	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.643	117	12915	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	2851	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	12742	132.39	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	264.78%#
35) Dibromofluoromethane	7.914	113	8159	82.62	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	165.24%#
50) Toluene-d8	10.344	98	15961	43.64	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	87.28%
62) 4-Bromofluorobenzene	12.626	95	4001	32.40	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	64.80%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD050321\
Data File : VD069049.D
Acq On : 03 May 2021 18:43
Operator : VA/SY
Sample : M2235-01
Misc : 4.54G/5.00ml/MSVOA_D/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
OR-02-043021

Quant Time: May 04 04:15:18 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D050321S.M
Quant Title : SW846 8260
QLast Update : Tue May 04 03:51:01 2021
Response via : Initial Calibration

