

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD050321\
 Data File : VD069053.D
 Acq On : 03 May 2021 20:40
 Operator : VA/SY
 Sample : M2263-02
 Misc : 5.05G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 S-1A

Quant Time: May 04 04:17:04 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D050321S.M
 Quant Title : SW846 8260
 QLast Update : Tue May 04 03:51:01 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.985	168	4010	50.00	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.861	114	7572	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.644	117	7624	50.00	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	3602	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	4870	96.16	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 192.32%#	
35) Dibromofluoromethane	7.914	113	2636	52.64	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 105.28%	
50) Toluene-d8	10.332	98	9072	48.92	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 97.84%	
62) 4-Bromofluorobenzene	12.620	95	3278	52.36	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 104.72%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD050321\
Data File : VD069053.D
Acq On : 03 May 2021 20:40
Operator : VA/SY
Sample : M2263-02
Misc : 5.05G/5.00ml/MSVOA_D/SOIL
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
S-1A

Quant Time: May 04 04:17:04 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D050321S.M
Quant Title : SW846 8260
QLast Update : Tue May 04 03:51:01 2021
Response via : Initial Calibration

