

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD050521\
 Data File : VD069069.D
 Acq On : 05 May 2021 15:09
 Operator : VA/SY
 Sample : M2265-01
 Misc : 7.08G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 124055

Quant Time: May 06 05:45:23 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D050421S.M
 Quant Title : SW846 8260
 QLast Update : Wed May 05 04:16:41 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.985	168	8940	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	15821	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.644	117	16136	50.00	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	5431	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	14251	123.69	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 247.38%#	
35) Dibromofluoromethane	7.920	113	8639	79.63	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 159.26%#	
50) Toluene-d8	10.344	98	20517	51.22	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 102.44%	
62) 4-Bromofluorobenzene	12.632	95	5793	43.97	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 87.94%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD050521\
Data File : VD069069.D
Acq On : 05 May 2021 15:09
Operator : VA/SY
Sample : M2265-01
Misc : 7.08G/5.00ml/MSVOA_D/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
124055

Quant Time: May 06 05:45:23 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D050421S.M
Quant Title : SW846 8260
QLast Update : Wed May 05 04:16:41 2021
Response via : Initial Calibration

