

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD050521\
 Data File : VD069071.D
 Acq On : 05 May 2021 16:07
 Operator : VA/SY
 Sample : M2296-01
 Misc : 7.24G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SP-1

Quant Time: May 06 05:45:58 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D050421S.M
 Quant Title : SW846 8260
 QLast Update : Wed May 05 04:16:41 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	31094	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	70601	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.644	117	81545	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	36494	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	50371	125.70	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery = 251.40%#			
35) Dibromofluoromethane	7.920	113	33899	70.02	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery = 140.04%			
50) Toluene-d8	10.338	98	88762	49.66	ug/l	0.00
Spiked Amount 50.000	Range 49 - 140		Recovery = 99.32%			
62) 4-Bromofluorobenzene	12.626	95	35418	60.24	ug/l	0.00
Spiked Amount 50.000	Range 25 - 144		Recovery = 120.48%			
Target Compounds						
20) Methylene Chloride	4.956	84	10368	22.70	ug/l	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD050521\
Data File : VD069071.D
Acq On : 05 May 2021 16:07
Operator : VA/SY
Sample : M2296-01
Misc : 7.24G/5.00ml/MSVOA_D/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SP-1

Quant Time: May 06 05:45:58 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D050421S.M
Quant Title : SW846 8260
QLast Update : Wed May 05 04:16:41 2021
Response via : Initial Calibration

