

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD050521\
 Data File : VD069072.D
 Acq On : 05 May 2021 16:36
 Operator : VA/SY
 Sample : M2279-01
 Misc : 6.33G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 EME-12

Quant Time: May 06 05:46:13 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D050421S.M
 Quant Title : SW846 8260
 QLast Update : Wed May 05 04:16:41 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	114308	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	222555	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.644	117	164319	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	39661	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	116332	78.97	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery = 157.94%			
35) Dibromofluoromethane	7.920	113	93998	61.59	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery = 123.18%			
50) Toluene-d8	10.338	98	246879	43.82	ug/l	0.00
Spiked Amount 50.000	Range 49 - 140		Recovery = 87.64%			
62) 4-Bromofluorobenzene	12.626	95	57313	30.92	ug/l	0.00
Spiked Amount 50.000	Range 25 - 144		Recovery = 61.84%			
Target Compounds					Qvalue	
16) Acetone	4.162	43	14620	50.96	ug/l	99
17) Carbon Disulfide	4.403	76	4900	1.21	ug/l	98
20) Methylene Chloride	4.967	84	29191	15.55	ug/l	91
25) 2-Butanone	7.214	43	6588	20.92	ug/l #	77
30) Chloroform	7.709	83	159176	58.48	ug/l	98
47) Bromodichloromethane	9.661	83	3533	1.47	ug/l #	72

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD050521\
Data File : VD069072.D
Acq On : 05 May 2021 16:36
Operator : VA/SY
Sample : M2279-01
Misc : 6.33G/5.00ml/MSVOA_D/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
EME-12

Quant Time: May 06 05:46:13 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D050421S.M
Quant Title : SW846 8260
QLast Update : Wed May 05 04:16:41 2021
Response via : Initial Calibration

