

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD050523\
 Data File : VD076019.D
 Acq On : 05 May 2023 12:07
 Operator : KP/SY
 Sample : 02591-06
 Misc : 5.01G/10.00ml/MSVOA_D/SOIL
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :

Quant Time: May 06 00:22:41 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\SFAMDLM042023SMA.M
 Quant Title : SFAM01.0
 QLast Update : Sat May 06 00:21:26 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	8.775	114	344463	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.581	117	316862	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.522	152	140691	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.269	65	176748	23.390	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	93.560%
7) Chloroethane-d5	2.799	69	155985	23.891	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	95.560%
11) 1,1-Dichloroethene-d2	3.916	65	41341	23.730	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	94.920%
21) 2-Butanone-d5	6.993	46	41516	63.286	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	126.580%
24) Chloroform-d	7.569	84	211529	24.557	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	98.240%
26) 1,2-Dichloroethane-d4	8.228	65	96246	26.165	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	104.640%
32) Benzene-d6	8.198	84	436717	23.647	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	94.600%
36) 1,2-Dichloropropane-d6	9.210	67	134512	26.314	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	105.240%
41) Toluene-d8	10.269	98	392327	22.691	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	90.760%
43) trans-1,3-Dichloroprop...	10.522	79	48998	23.190	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	92.760%
47) 2-Hexanone-d5	10.875	63	34358	55.822	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	111.640%
56) 1,1,2,2-Tetrachloroeth...	12.651	84	119266	30.490	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	121.960%#
66) 1,2-Dichlorobenzene-d4	13.816	152	124620	25.042	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	100.160%
Target Compounds						Qvalue
12) 1,1-Dichloroethene	3.922	96	1309	0.289	ug/L #	1
16) Methylene chloride	4.799	84	19816	2.836	ug/L	91
34) Trichloroethene	8.775	95	8659	1.747	ug/L #	2
35) Methylcyclohexane	9.210	83	31402	4.109	ug/L #	24
40) 4-Methyl-2-pentanone	10.269	43	1155	0.720	ug/L #	1


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD050523\
Data File : VD076019.D
Acq On : 05 May 2023 12:07
Operator : KP/SY
Sample : 02591-06
Misc : 5.01G/10.00ml/MSVOA_D/SOIL
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :

Quant Time: May 06 00:22:41 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\SFAMDLM042023SMA.M
Quant Title : SFAM01.0
QLast Update : Sat May 06 00:21:26 2023
Response via : Initial Calibration





#12

1,1-Dichloroethene

Concen: 0.289 ug/L

RT: 3.922 min Scan# 397

Delta R.T. -0.018 min


Lab File: VD076019.D

Acq: 05 May 2023 12:07

Instrument :

MSVOA_D

ClientSampleId :






Tgt Ion: 96 Resp: 1309

Ion Ratio Lower Upper

96 100

61 791.2 92.6 172.0#

63 6547.0 77.4 143.8#



#16

Methylene chloride

Concen: 2.836 ug/L

RT: 4.799 min Scan# 546

Delta R.T. 0.000 min

Lab File: VD076019.D

Acq: 05 May 2023 12:07




Tgt Ion: 84 Resp: 19816


Ion Ratio Lower Upper

84 100

86 66.0 45.8 85.2

49 101.4 60.8 112.8









#34
Trichloroethene
Concen: 1.747 ug/L
RT: 8.775 min Scan# 11
Delta R.T. -0.253 min
Lab File: VD076019.D
Acq: 05 May 2023 12:07

Instrument :
MSVOA_D
ClientSampleId :

Tgt Ion: 95 Resp: 8659




Ion	Ratio	Lower	Upper
95	100		
97	0.0	46.0	85.4#
132	0.0	73.3	136.1#
130	0.0	75.3	139.9#




#35
Methylcyclohexane
Concen: 4.109 ug/L
RT: 9.210 min Scan# 1296
Delta R.T. -0.065 min
Lab File: VD076019.D
Acq: 05 May 2023 12:07

Tgt Ion: 83 Resp: 31402


Ion	Ratio	Lower	Upper
83	100		
55	0.4	48.2	72.2#
98	0.9	40.8	61.2#





#40
 4-Methyl-2-pentanone
 Concen: 0.720 ug/L
 RT: 10.269 min Scan# 1457
 Delta R.T. 0.112 min
 Lab File: VD076019.D
 Acq: 05 May 2023 12:07

Instrument :
 MSVOA_D
 ClientSampleId :



Tgt Ion: 43 Resp: 1155
 Ion Ratio Lower Upper
 43 100
 58 235.7 39.1 58.7#
 100 21648.7 19.4 29.0#

